

EU Type Examination Certificate

This is to certify that:

GVS Filter Technology UK Ltd
NFC House
Vickers Industrial Estate
Mellishaw Lane
Morecambe
Lancashire
LA3 3EN
United Kingdom

Holds Certificate Number:


CE 711763

In respect of:

Filtering masks to EN 149:2001+A1:2009
See Continuation sheets for details.

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified
Body for the above Regulation
(Notified Body Number 2797):



Drs. Dave Hagenaaars, Managing Director

First Issued: 2019-05-01
Latest Issue: 2022-06-17

Effective Date: 2022-06-17
Expiry Date: 2024-05-01

Page: 1 of 4



...making excellence a habit.™

EU Type Examination Certificate

No. CE 711763

Product Specification

Vertical Fold Flat Range with clogging

Description: Vertical fold flat disposable face mask with clogging option 7.17d

Technical specification: **EN 149:2001+A1:2009** Respiratory protective devices – Filtering half masks to protect against particles.

Classification – Reusable – R or Single shift - NR

FFP Class: **FFP1 R D**

Model:
CN P1 F10000
CN P1V F10050 (mask with exhalation valve fitted)

FFP Class: **FFP1 NR D**

Model:
CN P1C F10005 (mask with activated charcoal)
CNP1CV F10055 (mask with activated charcoal and exhalation valve fitted)

FFP Class: **FFP2 R D**

Model:
CN P2 F20000
CN P2V F20050 (mask with exhalation valve fitted)

FFP Class: **FFP2 NR D**

Model: CN P2CV F20055 (mask with activated charcoal and exhalation valve fitted)

FFP Class: **FFP2 NR**

CN P2 F21000

First Issued: 2019-05-01

Latest Issue: 2022-06-17

Effective Date: 2022-06-17

Expiry Date: 2024-05-01

Page: 2 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 711763

Product Specification

Description: Vertical fold flat disposable face mask with clogging option 7.17d

Technical specification: EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles.

Classification – Reusable – R or Single shift - NR

FFP Class: FFP3 R D

Model: CN P3 F30000

CN P3V F30050 (mask with exhalation valve fitted)

CN P3 F31000 Classified as a Surgical Mask Type IIR EN 14683 for use as a Medical Device - Class 1

FFP Class: FFP3 NR D

Model: CN P3CV F30055 (mask with activated charcoal and exhalation valve fitted)

FFP Class FFP3 NR

CN P3 F31000 Classified as a Surgical Mask Type IIR EN 14683 for use as a Medical Device - Class 1

CN P3 F31050 (mask with exhalation valve fitted)

First Issued: 2019-05-01

Latest Issue: 2022-06-17

Effective Date: 2022-06-17

Expiry Date: 2024-05-01

Page: 3 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 711763

Certificate Administration Details

Technical File Reference: Particle Filtering Half Masks EN149 Cap.01/Rev. 11

Certificate Amendment Record

Issue date	Comments	BSI Internal Project No.
June 2022	Addition of F31050 and F31000 NR	2797:22:3617508
January 2022	Addition of F21000	2797:22:3585600
December 2021	Administration update of F31000 to include the medical device classification.	2797:21:3555181
December 2020	Update of the Monitoring of manufactured PPE details.	2797:20:3205346
June 2020	Addition of model F31000 and Module C2 reference for GVS Italy, Bologna, Sud and GVS China, Suzhou and GVS Romania, Ciorani. Italcert C2 up-issued to PPE194AC379.	2797:20:3174968
April 2020	Addition of Italcert Module C2 reference for GVS Italy, Bologna.	2797:20:3166806
May 2019	First issue. Transition from certificate CE 672393 to the PPE Directive 89/686.	0086:19:9690679

Note: The Certificate holder is responsible for keeping the Notified Body advised of changes to any aspect of the overall process used in the manufacture of the product.

Monitoring of manufactured PPE.

The EC Quality of Production by Monitoring requirements, Module D, for the product are referenced in BSI issued Certificate Number CE 672394.

First Issued: 2019-05-01

Latest Issue: 2022-06-17

Effective Date: 2022-06-17

Expiry Date: 2024-05-01

Page: 4 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.