

# PRODUCT SPECIFICATION

Product P/N	2800/22	Mod. 984A Rev. 06
Description	Spiroguard	

**2800/22**

**Spiroguard**



<b>PRODUCT DESCRIPTION</b>	<p>Inlet Outlet Connectors:          OD34mm ID29.25mm - Machine Side;          OD31.2mm ID26.09mm - Patient Side.          Approx. Dimensions: 97mm diameter x 85mm height.          Weight: 37gm (approx.).          Bidirectional Filter.</p>
<b>MANUFACTURER NAME</b>	<p><b>GVS Filter Technology UK</b>          NFC House          Vickers Industrial Estate          Mellishaw Lane, Morecambe          Lancashire LA3 3EN - United Kingdom</p> <p><b>Information</b>          Tel. +44 (0) 1524 847600          e-mail: gvsuk@gvs.com</p>
<b>INTENDED USE / APPLICATION</b>	<p>Spirometer pre filters are used in pulmonary function tests for trapping bacteria and viruses and other particulates, reducing the risk of cross-contamination between the patient and the machine. Filters can be supplied with accessories, such as nose clip, mouthpiece and bite grip.</p>
<b>CLASS OF THE PRODUCT</b>	<p>Disposable medical device - Class IIa          Rule 2 Annex IX 93/42 / EEC          Rule 5 Annex VIII MDR 2017/745</p>
<b>MATERIALS</b>	<p><b>Filter media: <i>Electrostatic Blended Synthetic Fiber</i></b>  <b>Frame/Housing Polymer: <i>White Polypropylene Homopolymer (PP)</i></b>  <b>Colour: <i>White</i>.</b></p> <p><b>Regulatory Documentation Required:</b></p> <ul style="list-style-type: none"> <li>- Biocompatibility according ISO 10993-1</li> <li>- ROHS</li> <li>- BSE/TSE</li> <li>- DEHP plasticizer Free and latex free</li> <li>- Aging</li> <li>- REACH</li> <li>- Conflict minerals</li> </ul>

# PRODUCT SPECIFICATION

<b>Product P/N</b>	<b>2800/22</b>	<b>Mod. 984A Rev. 06</b>
<b>Description</b>	<b>Spiroguard</b>	

## PRODUCT CHARACTERISTICS

### Appearance/Visual

As shown on drawing.

### Physical/Mechanical

Approx. Dimensions: **97mm diameter x 85mm height.**

Weight: **37gm (approx.).**

Interfaces (ex: Input / Output connectors): **OD34mm ID29.25mm - Machine Side;**  
**OD31.2mm ID26.09mm - Patient Side.**

Operating temperature Range: **N/A**

Storage temperature Range: **5 °C to 40 °C**

**Bidirectional Filter, Male connector – Patient Side.**

### Biological

Pyrogenicity: **<0.3 EU/ml**

**Biocompatibility to ISO10993**

Category – **Surface device**

Contact – **Oral cavity**

Contact Duration - **<24hrs**

### Functional

Air Flow Rate: **30l/min, 60l/min, 90l/min.**

Filtration Efficiency: **Filter Efficiency @ 30L/min using TSI 8130: Min. 99%**  
(REP: 1433/17 with factor of safety applied to Min.)

Pressure Drop:

**Flow Resistance @ 30l/min in accordance with EN ISO 9360-1: Max.35.2Pa**

**Flow Resistance @ 60l/min in accordance with EN ISO 9360-1: Max.75.9Pa**

**Flow Resistance @ 90l/min in accordance with EN ISO 9360-1: Max.132Pa**

(REP:1255/17 with 10% of safety margin added to Max.)

Internal Volume: **100ml (approx.)**

Operating Lifetime: **Refer to Instructions for Use.**

Shelf Lifetime: **5 years from the date of manufacture.**

**Bacterial Filtration Efficiency in accordance with ASTM F2101-07: Min. 99.999%**  
(*Staphylococcus aureus* @ 30L /minute) REP: EXT607770.

**Viral Filtration Efficiency in accordance with ASTM F2101-07: Min. 99.999%**  
(*Bacteriophage* @ 30L /minute) REP: EXT620332.

### Cleanliness

Device assembled within Class 8 Cleanroom.

### Testing

**Torque test @ 10Nm.** (REP: 1348/17)


# PRODUCT SPECIFICATION

<b>Product P/N</b>	<b>2800/22</b>	<b>Mod. 984A</b> <b>Rev. 06</b>
<b>Description</b>	<b>Spiroguard</b>	
<b>INSTRUCTIONS / WARNINGS</b>	<i>Multi-language IFU available.</i>	
<b>PRODUCT SHELF LIFE</b>	<i>5 years from the date of manufacture.</i> <i>Expiration date and date of manufacture are detailed on the product labelling.</i>	
<b>APPLICABLE STANDARDS AND REGULATIONS</b>	<p><b>Product Certification required:</b></p> <ul style="list-style-type: none"> <li>- CE mark</li> <li>- FDA</li> </ul> <p><b>Applicable Standards and Technical Regulations:</b></p> <p><i>Biological evaluation of Medical Devices - Part 1: Evaluation and Testing - ISO 10993-1.</i></p> <p><i>Respiratory protective devices - Method for test - Part 7: Determination of particle filter penetration - BS EN 13274-7.</i></p> <p><i>Medical devices- Application of risk management to medical devices - BS EN ISO 14971.</i></p> <p><i>Medical devices – symbols to be used with medical device labels, labelling and information to be supplied - Part1: General requirements - ISO 15223-1.</i></p>	
<b>PACKAGING AND LABELING</b>	<p><i>Number of pcs per bag is determined by the sales order.</i></p> <p><i>The first barcode label is applied to the outside of the bags.</i></p> <p><i>The second barcode label is applied onto the outside of the box.</i></p> <p><i>Each bag is labelled with the following traceability information:</i></p> <ul style="list-style-type: none"> <li>✓ Quantity</li> <li>✓ Product description</li> <li>✓ Product Date</li> <li>✓ Lot Number (OL and 5-digit batch number to trace back to raw materials used)</li> <li>✓ Operator Code</li> </ul> <p><i>Different lots in one box are separately closed and separately labelled.</i></p> <p><i>Bulk products will be packed in double PE bags.</i></p>	
<b>CERTIFICATE OF COMPLIANCE</b>	<p><i>With each shipment, GVS UK Customer Service will send the CofC to the Customer, based on the lot numbers and date of manufacture.</i></p> <p><i>Conformity declaration is printed on every invoice and Certificate is according to UNI EN 10204 type 2.1.</i></p> <p><i>The Quality management system is in compliance with ISO 9001, ISO 13485.</i></p>	
<b>DRAWING</b>	<i>The attached drawing is part of this product specification and must not be duplicated or made accessible to a third party without written permission from GVS Filter Technology UK Ltd.</i>	



# PRODUCT SPECIFICATION

Product P/N	2800/22	Mod. 984A Rev. 06
Description	Spiroguard	



Approximate dimensions for reference only

## ACCEPTABLE QUALITY LEVEL

AQL: 0.65 with sampling Plan: ISO2859.

## VISUAL REQUIREMENTS

**Visual acceptance requirements apply when inspected under below conditions:**

Magnification: *Unaided eye at a distance of approximately 35-40cm.*  
 Light type: *Lighting level must be reasonable for visual detection.*  
 Timings: *Maximum inspection period per item is 25 seconds.*  
 For detailed defect list, refer to product control plan.

Acceptance Requirement		AQL	Sampling Plan
1	Black particle contamination	0.65	ISO 2859 Part 1 General Inspection Level 1
2	Damaged/broken item	0.65	
3	Blocked connector/luer	0.65	
4	Weld marks	0.65	
5	Short fill moulding	0.65	
6	Rough surface or edges	0.65	

# PRODUCT SPECIFICATION

<b>Product P/N</b>	<b>2800/22</b>	<b>Mod. 984A Rev. 06</b>
<b>Description</b>	<b>Spiroguard</b>	

	7	Pronounced injection gate	0.65
	8	Deformation/distortion	0.65
	9	Crack	0.65
	10	Oil/grease	0.65
	11	Wrong colour	0.65
	12	Weld fault	0.65

**GENERAL SAFETY AND PERFORMANCE REQUIREMENTS**

**Special characteristic:** Product characteristic which can affect safety or compliance with regulations, fit, function, performance or subsequent processing of product.

**Special Characteristic # 01:**

*Flow Resistance @ 30L/min in accordance with EN ISO 9360-1*

*Flow Resistance @ 60L/min in accordance with EN ISO 9360-1*

*Flow Resistance @ 90L/min in accordance with EN ISO 9360-1*

**Special Characteristic # 02:** Filter Efficiency @ 30L/min using TSI 8130 in accordance with EN 13274-7.

**Special Characteristic # 03:** Bacterial Filtration Efficiency in accordance with ASTM F2101-07

*Viral Filtration Efficiency in accordance with ASTM F2101-07*

This material specification describes the properties of product above indicated. This document contains general requirements, material description, drawing references, defect specification, biological material requirements.

**REVISIONS AND APPROVALS:**

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (NAME/FUNCTION/SIGNATURE)	APPROVED BY: (NAME/FUNCTION/SIGNATURE)
10/08/2020	2	Filter Efficiency updated.	Kinga Gawdzik – Engineering Support Technician 	Andrew Pearce – Quality Manager 



# PRODUCT SPECIFICATION

Product P/N	2800/22	Mod. 984A Rev. 06
Description	Spiroguard	

## CUSTOMER APPROVAL:

We accept this material specification as a part of the agreed terms of delivery.

Company Name: \_\_\_\_\_

Approved by: \_\_\_\_\_

NAME/FUNCTION

SIGNATURE

DATE

COMPANY STAMP

*Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.*