


PRODUCT SPECIFICATION

Product P/N	2200/05	Mod. 984A Rev. 06
Description	Insufflation Filter	

2200/05

Insufflation Filter



PRODUCT DESCRIPTION	Inlet/Outlet Connectors: 5mm - 9.5mm hose barbed - both sides. Approx. dimensions: 65mm diameter x 65.8mm height. Bidirectional filter. Weight 17gm (approx.).
MANUFACTURER NAME	GVS Filter Technology UK NFC House Vickers Industrial Estate Mellishaw Lane, Morecambe Lancashire LA3 3EN - United Kingdom Information Tel. +44 (0) 1524 847600 e-mail: gvsuk@gvs.com
INTENDED USE / APPLICATION	To be used in conjunction with insufflators for intra-abdominal insufflation.
CLASS OF THE PRODUCT	Disposable medical device - Class IIa Rule 2 Annex IX 93/42 / EEC Rule 2 Annex VIII MDR 2017/745
MATERIALS	Filter media: <i>Hydrophobic Glass Microfibre Media</i> Frame/Housing Polymer: <i>Clear Styrene-Butadiene Copolymer (SBC)</i> Colour: <i>Transparent Clear</i> Regulatory Documentation Required: <ul style="list-style-type: none"> - Biocompatibility according ISO 10993-1 - ROHS - BSE/TSE - DEHP plasticizer Free and latex free - Aging - REACH - Conflict minerals
PRODUCT CHARACTERISTICS	Appearance/Visual As shown on drawing.


PRODUCT SPECIFICATION

Product P/N	2200/05	Mod. 984A Rev. 06
Description	Insufflation Filter	

	<p>Physical/Mechanical <i>Approx. dimensions: 65mm diameter x 65.8mm height.</i> <i>Weight: 17gm (approx.).</i> Interfaces (ex: Input / Output connectors): 5mm - 9.5mm hose barbed - both sides. Operating temperature Range: N/A Storage temperature Range: 5 °C to 40 °C Bidirectional Filter.</p> <p>Biological Pyrogenicity: <0.3 EU/ml Biocompatibility to ISO10993 Category – Surface device Contact – Skin Contact Duration - <24hrs</p> <p>Functional Air Flow Rate: Min. 28.4l/min @ 5PSI (REP 0614/15 with 20% Factor of Safety applied to Min.)</p> <p>Filtration Efficiency: <i>Filter Efficiency @ 15L/min using DEHS@0.3µm: Min.99.9% (Ref. 2200/15 REP: 1218/17 with Factor of Safety)</i> Pressure Drop: N/A Internal Volume: N/A Operating Lifetime: Refer to Instructions for Use. Shelf Lifetime: 5 years from the date of manufacture.</p> <p><i>Bacterial Filtration Efficiency in accordance with ASTM F2101-07: Min. 99.999% (Staphylococcus aureus @ 30L /minute) REP: EXT846022-S01</i></p> <p><i>Viral Filtration Efficiency in accordance with ASTM F2101-07: Min. 99.999% (Bacteriophage @ 30L/ minute) REP: EXT846021-S01</i></p> <p>Cleanliness Device assembled within Class 8 Cleanroom.</p> <p>Testing Burst test: Min.15.7psi (REP 0618/15 with 20% Factor of Safety applied to Minimum) Leakage: Static Head Test 27" of water for 1 minute (REP 0622/15)</p>
INSTRUCTIONS / WARNINGS	Multi-language IFU available.
PRODUCT SHELF LIFE	5 years from the date of manufacture. <i>Expiration date and date of manufacture are detailed on the product labelling.</i>
STERILIZATION	Sterile version of product available (Ethylene oxide - Max 55°C)
APPLICABLE STANDARDS AND REGULATIONS	<p>Product Certification required:</p> <ul style="list-style-type: none"> - CE mark - FDA <p>Applicable Standards and Technical Regulations:</p>

PRODUCT SPECIFICATION

Product P/N	2200/05	Mod. 984A
Description	Insufflation Filter	Rev. 06

	<p><i>Biological evaluation of Medical Devices - Part 1: Evaluation and Testing - ISO 10993-1.</i></p> <p><i>Medical devices- Application of risk management to medical devices - BS EN ISO 14971.</i></p> <p><i>Medical devices – symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements - ISO 15223-1.</i></p> <p><i>Sterilization of health care products – Ethylene oxide sterilization – ISO 11135-1.</i></p> <p><i>Sterilization of medical devices – Microbiological Methods – Part 1: Estimation of population of Microorganisms on products – ISO 11737-1.</i></p> <p><i>High Efficiency Air Filters – BS EN 1822.</i></p>
PACKAGING AND LABELING	<p><i>Number of pcs per bag is determined by the sales order.</i></p> <p><i>The first barcode label is applied to the outside of the bags.</i></p> <p><i>The second barcode label is applied onto the outside of the box.</i></p> <p><i>Each bag is labelled with the following traceability information:</i></p> <ul style="list-style-type: none"> ✓ Quantity ✓ Product description ✓ Product Date ✓ Lot Number (OL and 5-digit batch number to trace back to raw materials used) ✓ Operator Code <p><i>Different lots in one box are separately closed and separately labelled.</i></p> <p><i>Bulk products will be packaged in double PE bags</i></p>
CERTIFICATE OF COMPLIANCE	<p><i>With each shipment, GVS UK Customer Service will send the CoFC to the Customer, based on the lot numbers and date of manufacture.</i></p> <p><i>Conformity declaration is printed on every invoice and Certificate is according to UNI EN 10204 type 2.1.</i></p> <p><i>The Quality management system is in compliance with ISO 9001, ISO 13485.</i></p>
DRAWING	<p><i>The attached drawing is part of this product specification and must not be duplicated or made accessible to a third party without written permission from GVS Filter Technology UK Ltd.</i></p> <div style="text-align: center;">  <p>Top view: Circular filter with outer diameter $\varnothing 65$ and inner diameter $\varnothing 19$. The text 'MI' is on the left and 'IN' is at the bottom.</p> <p>Side view: Shows the filter's profile with a total height of 65.8.</p> </div> <div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: fit-content;"> <p>Approximate dimensions for reference only</p> </div>

PRODUCT SPECIFICATION

Product P/N	2200/05	Mod. 984A Rev. 06
Description	Insufflation Filter	



ACCEPTABLE QUALITY LEVEL	AQL: with sampling Plan: 0.65 AQL to ISO 2859-1			
VISUAL REQUIREMENTS	Visual acceptance requirements apply when inspected under below conditions: Magnification: <i>Unaided eye at a distance of approximately 35-40cm.</i> Light type: <i>Lighting level must be reasonable for visual detection.</i> Timings: <i>Maximum inspection period per item is 25 seconds.</i> <i>For detailed defect list, refer to product control plan.</i>			
	Acceptance Requirement		AQL	Sampling Plan
	1	Black particle contamination	0.65	ISO 2859 Part 1 General Inspection Level 1
	2	Damaged/broken item	0.65	
	3	Blocked connector/luer	0.65	
	4	Short fill moulding	0.65	
	5	Rough surface or edges	0.65	
	6	Pronounced injection gate	0.65	
	7	Deformation/distortion	0.65	
	8	Crack	0.65	
	9	Oil/grease	0.65	
	10	Wrong colour	0.65	
	11	Weld fault	0.65	
	12	Weld marks	0.65	
GENERAL SAFETY AND PERFORMANCE REQUIREMENTS	Special characteristic: <i>Product characteristic which can affect safety or compliance with regulations, fit, function, performance or subsequent processing of product.</i>			
	Special Characteristic # 01: <i>Air Flow rate at 5psi</i>			
	Special Characteristic # 02: <i>Filter Efficiency @ 15L/min using DEHS @ 0.3µm in accordance with High Efficiency Air Filters – BS EN 1822.</i>			
	Special Characteristic # 03: <i>Bacterial Filtration Efficiency in accordance with ASTM F2101-07</i> <i>Viral Filtration Efficiency in accordance with ASTM F2101-07</i>			
	Special Characteristic # 04: <i>Burst Pressure</i>			
	Special Characteristic # 05: <i>27” water leak test – Static head test</i>			

PRODUCT SPECIFICATION

Product P/N	2200/05	Mod. 984A
Description	Insufflation Filter	Rev. 06

This material specification describes the properties of product above indicated. This document contains general requirements, material description, drawing references, defect specification, biological material requirements.

REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (NAME/FUNCTION/SIGNATURE)	APPROVED BY: (NAME/FUNCTION/SIGNATURE)
27/08/2021	2	Staerilisation section and functional characteristics amended.	Kinga Gawdzik - Engineering Support Technician 	Andrew Pearce – Quality Manager 

CUSTOMER APPROVAL:

We accept this material specification as a part of the agreed terms of delivery.

Company Name:

Approved by:

NAME/FUNCTION

SIGNATURE

DATE

COMPANY STAMP

Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.