

Product P/N	4222/702	
		Mod. 984A
Description	MAXI Maxi Filter	
Description		NEV. 00

4222/702

MAXI Maxi Filter



PRODUCT DESCRIPTION	Inlet/Outlet Connectors: 22mm Male/15mm Female – 22mm Female/15mm Male ISO. Approx. dimensions: 68.5mm diameter x 67.1mm height. Weight: 25g (approx.). Bidirectional Filter.
MANUFACTURER NAME	GVS Filter Technology UK NFC House Vickers Industrial Estate Mellishaw Lane, Morecambe Lancashire LA3 3EN - United Kingdom Information Tel. +44 (0) 1524 847600 e-mail: gvsuk@gvs.com
INTENDED USE / APPLICATION	Filters protect the patient's airways effectively from exogenous microbial loads, thus reducing the risk of extrinsic colonisation and infection. Used to help reduce cross contamination between patient and machine.
CLASS OF THE PRODUCT	Disposable medical device - Class IIa Rule 2 Annex IX 93/42 / EEC Rule 5 Annex VIII MDR 2017/745
MATERIALS	Filter media: Electrostatic Blended Synthetic Fiber         Frame/Housing Polymer: Transparent Green Tinted Polypropylene (PP)         Colour: Transparent Green         Regulatory Documentation Required:         -       Biocompatibility according ISO 10993-1         -       ROHS         -       BSE/TSE         -       DEHP plasticizer Free and latex free         -       Aging         -       REACH         -       Conflict minerals
PRODUCT CHARACTERISTICS	Appearance/Visual As shown on drawing.



Product P/N	4222/702 Mod. 984A	
Description	MAXI Maxi Filter Rev. 06	•
	Physical/Mechanical         Approx. dimensions: 68.5mm diameter x 67.1mm height.         Weight: 25g (approx.).         Interfaces (ex: Input / Output connectors): 22mm Male/15mm Female – 22mm Female/15mm         Male ISO.         Operating temperature Range: N/A	m
	Storage temperature Range: 5 °C to 40 °C Bidirectional Filter.	
	Biological Pyrogenicity: <0.3 EU/ml Biocompatibility to ISO10993 Category – Surface device Contact – Skin Contact Duration - <24hrs	
	Functional Air Flow Rate: <b>301/min, 601/min, 901/min.</b>	
	Filtration Efficiency: <i>NaCl Filter Efficiency</i> @ 30L/min using TSI 8130: <b>Min. 98.5%</b> (REP: 0840/16 with factor of safety applied to Min.)	
	Pressure Drop: Flow Resistance @ 30l/min in accordance with EN ISO 9360-1: Max. 99Pa Flow Resistance @ 60l/min in accordance with EN ISO 9360-1: Max. 220Pa Flow Resistance @ 90l/min in accordance with EN ISO 9360-1: Max. 363Pa (REP:0783/16 with 10% of safety margin added to Max.)	
	Internal Volume: 29ml (approx.)	
	Operating Lifetime: Refer to Instructions for Use.	
	Shelf Lifetime: 5 years from the date of manufacture.	
	Bacterial Filtration Efficiency in accordance with ASTM F2101-07: <b>Min. 99.999%</b> (Staphylococcus aureus @ 30L /minute) Ref.4222/701 REP: EXT486704B	
	Viral Filtration Efficiency in accordance with ASTM F2101-07: <b>Min. 99.999%</b> (Bacteriophage @ 30L/ minute) Ref. 4222/701 REP: EXT486705B.1	
	Gas leakage in accordance with EN9360: Max.0.0 ml/min (REP:2331/21)	
	<b>Cleanliness</b> Device assembled within Class 8 Cleanroom.	
	Testing Leak test at 3PSI.	
INSTRUCTIONS / WARNINGS	Multi-language IFU available.	



Product P/N	4222/702	
	Mod. 984A	
Description	MAXI Maxi Filter Rev. 06	
PRODUCT SHELF	5 years from the date of manufacture.	
	Expiration date and date of manufacture are detailed on the product labelling.	
STERILIZATION	Sterile version of product available (Ethylene oxide - Max 55°C)	
APPLICABLE STANDARDS AND REGULATIONS	Product Certification required: - CE mark - FDA	
	<b>Applicable Standards and Technical Regulations:</b> Biological evaluation of Medical Devices - Part 1: Evaluation and Testing - ISO 10993-1.	
	Respiratory protective devices - Method for test - Part 7: Determination of particle filter penetration - BS EN 13274-7.	
	Medical devices- Application of risk management to medical devices - BS EN ISO 14971.	
	Medical devices – symbols to be used with medical device labels, labelling and information to a supplied - Part1: General requirements - ISO 15223-1.	
	Breathing system filters for anaesthetic and respiratory use — Part 1: Salt test method to asse filtration performance - ISO 23328-1.	
	Anaesthetic and respiratory equipment – conical connectors – part 1: Cones and sockets – ISO 5356-1.	
	Sterilization of health care products – Ethylene oxide sterilization – ISO 11135-1.	
	Sterilization of medical devices – Microbiological Methods – Part 1: Estimation of population of microorganisms on products – ISO 11737-1.	
PACKAGING AND		

PACKAGING AND	
LABELING	Number of pcs per bag is determined by the sales order. The first barcode label is applied to the outside of the bags. The second barcode label is applied onto the outside of the box. Each bag is labelled with the following traceability information: <ul> <li>Quantity</li> <li>Product description</li> <li>Product Date</li> <li>Lot Number (OL and 5-digit batch number to trace back to raw materials used)</li> <li>Operator Code Different lots in one box are separately closed and separately labelled. Bulk products will be packed in double PE bags.</li> </ul>
CERTIFICATE OF COMPLIANCE	With each shipment, GVS UK Customer Service will send the CofC to the Customer, based on the lot numbers and date of manufacture. Conformity declaration is printed on every invoice and Certificate is according to UNI EN 10204 type 2.1. The Quality management system is in compliance with ISO 13485.
DRAWING	The attached drawing is part of this product specification and must not be duplicated or made accessible to a third party without written permission from GVS Filter Technology UK Ltd.



Product P/N	4222/702	
		Mod. 984A
Description	MAXI Maxi Filter	Rev. 06

	22 Female / 15 Male Connector in accordance with EN ISO 5356 22 Male / 15 Female Connector in accordance with EN ISO 5356		
ACCEPTABLE	AQL: 0.65 with sampling Plan: ISO2859.		
QUALITY LEVEL VISUAL REQUIREMENTS	Visual acceptance requirements apply when inspected under be Magnification: Unaided eye at a distance of approximately 35-40cm. Light type: Lighting level must be reasonable for visual detection. Timings: Maximum inspection period per item is 25 seconds. For detailed defect list, refer to product control plan.		
	Acceptance Requirement	AQL	Sampling Plan
	1 Black particle contamination	0.65	
	2 Damaged/broken item	0.65	ISO 2859 Part 1 General Inspection
	3 Blocked connector/luer	0.65	Level 1
	4 Weld marks	0.65	



Product P/N	4222/702	Mod. 984A
Description	MAXI Maxi Filter	Rev. 06

	5	Short fill moulding	0.65		
	6	Rough surface or edges	0.65		
	7	Pronounced injection gate	0.65		
	8	Deformation/distortion	0.65		
	9	Crack	0.65		
	10	0 Oil/grease	0.65		
	11	Wrong colour	0.65		
	12	Weld fault	0.65		
GENERAL SAFETY AND PERFORMANCE REQUIREMENTS	Special characteristic: Product characteristic which can affect safety or compliance with regulations, fit, function, performance or subsequent processing of product. Special Characteristic # 01:				
	Flow	Resistance @ 30L/min in accordance with EN ISO 9360-1,			
	Flow	Resistance @ 60L/min in accordance with EN ISO 9360-1,			
	Flow	Resistance @ 90L/min in accordance with EN ISO 9360-1.			
	<b>Special Characteristic # 02:</b> NaCl Filter Efficiency @ 30L/min using TSI 8130 in accordance with EN 13274-7.				
	Speci	al Characteristic # 03: Bacterial Filtration Efficiency in accord	dance with ASTM F2101-07,		
	Viral I	Filtration Efficiency in accordance with ASTM F2101-07.			
	Speci	al Characteristic # 04: Conical connectors compliant in acco	rdance with EN5356.		
	Special Characteristic # 05: Gas Leakage compliant in accordance with EN9360.				
			This material specification describes the properties of product above indicated. This document contains general requirements, material description, drawing references, defect specification, biological material requirements.		

#### **REVISIONS AND APPROVALS:**

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (NAME/FUNCTION/SIGNATURE)	APPROVED BY: (NAME/FUNCTION/SIGNATURE)
03/08/2021	4	Functional characteristics updated.	Kinga Gawdzik – Engineering Support Technician	Andrew Pearce – Quality Manager



Product P/N	4222/702	Mod. 984A
Description	MAXI Maxi Filter	Rev. 06

CUSTOMER APPROVAL:		
We accept this	material specification as a part of the agreed terms of delivery.	
	Company Name:	
Approved by:		
-	NAME/FUNCTION	
-	SIGNATURE	
	DATE	
-	COMPANY STAMP	

Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.