

Product P/N	2802/30	Mod. 984A
Description	Spiroguard adaptor	

2802/30

Spiroguard adaptor



PRODUCT DESCRIPTION	Inlet/Outlet connectors: OD30mm/ID28mm and OD35.2mm/28.7mm (approx.) Approx. Dimensions: 20mm length. Weight: 15g (approx.)
MANUFACTURER NAME	GVS Filter Technology UK NFC House Vickers Industrial Estate Mellishaw Lane, Morecambe Lancashire LA3 3EN - United Kingdom Information Tel. +44 (0) 1524 847600 e-mail: gvsuk@gvs.com
INTENDED USE / APPLICATION	Spiroguard Machine Adaptor acts as an adaptor between the spiroguard filter and the pulmonary function machine.
CLASS OF THE PRODUCT	Disposable medical device - Class IIa Rule 2 Annex IX 93/42/EEC Rule 2 Annex VIII MDR 2017/745
MATERIALS	Filter media: N/A Frame/Housing Polymer: Acetal Colour: White Regulatory Documentation Required: - ROHS - BSE/TSE - DEHP plasticizer Free and latex free - Aging - REACH - Conflict minerals
PRODUCT CHARACTERISTICS	Appearance/Visual As shown on drawing



Product P/N	2802/30 Mo	d. 984A	
Description		_ Mod. 984A Rev. 06	
	Physical/Mechanical		
	Approx. Dimensions: 20mm length. Weight: 15g (approx.) Interfaces (ex: Input / Output connectors):		
	OD30mm/ID28mm and OD35.2mm/28.7mm (approx.) Operating temperature Range: N/A Storage temperature Range: 5 °C to 40 °C		
	Biological Pyrogenicity: <0.3 EU/mI Category – Surface device Contact – Skin Contact Duration - <24hrs		
	Functional Air Flow Rate: N/A Filtration Efficiency: N/A Pressure Drop: N/A Internal Volume: N/A Operating Lifetime: Refer to Instructions for Use.		
	Shelf Lifetime: 5 years from the date of manufacture. Cleanliness Part cleanliness: Hand wash adaptor with hot water and a mild soap. Rinse it water, shake off excess water and allow to air dry. Wipe with alcohol impres- before reconnecting it to the machine. Adaptor must be cleaned after each p	nated wipes	
	Testing N/A		
NSTRUCTIONS / WARNINGS	Multi-language IFU available.		
PRODUCT SHELF L	IFE 5 years from the date of manufacture.		
TERILIZATION	Expiration date and date of manufacture are detailed on the product labelling.		
PPLICABLE	Product Certification required:		
STANDARDS AND REGULATIONS	- CE mark - FDA		
	Applicable Standards and Technical Regulations: Medical devices- Application of risk management to medical devices - BS EN ISC) 14971.	
	Medical devices – symbols to be used with medical device labels, labelling and in be supplied - Part1: General requirements - ISO 15223-1.	formation to	



Product P/N	2802/30 Mod. 984/	Δ	
Description	Spiroguard adaptor Rev. 06		
PACKAGING AND LABELING	Number of pcs per bag is determined by the sales order. The first barcode label is applied to the outside of the bags. The second barcode label is applied onto the outside of the box. Each bag is labelled with the following traceability information: \checkmark Quantity \checkmark Product description \checkmark Product Date \checkmark Lot Number (OL and 5-digit batch number to trace back to raw materials \checkmark Operator Code Different lots in one box are separately closed and separately labelled. Bulk products will be packed in double PE bags.	used)	
CERTIFICATE OF COMPLIANCE	 With each shipment, GVS UK Customer Service will send the CofC to the Customer, bas the lot numbers and date of manufacture. Conformity declaration is printed on every invoice and Certificate is according to UNI EN type 2.1. The Quality management system is in compliance with ISO 13485. 		
DRAWING	The attached drawing is part of this product specification and must not be duplicated or accessible to a third party without written permission from GVS Filter Technology UK. Approximate dimensions for reference only $ \frac{\phi_{35.2}}{\phi_{30}} \qquad $		
ACCEPTABLE QUALITY	Y AQL: 0.65 with sampling Plan: ISO2859.		



Product P/N	roduct P/N 2802/30				
Description	Spire	ogua	Rev. 06		
VISUAL REQUIREMENTS	N Li T	lagnif ight ty iming	acceptance requirements apply when inspected under ication: Unaided eye at a distance of approximately 35-40c upe: Lighting level must be reasonable for visual detection. s: Maximum inspection period per item is 25 seconds. tailed defect list, refer to product control plan.		nditions:
			Acceptance Requirement	AQL	Sampling Plan
		1	Black particle contamination	0.65	
		2	Damaged/broken item	0.65	
		3	Rough surface or edges	0.65	ISO 2859 Part 1
		4	Deformation/distortion	0.65	General Inspection
		5	Crack	0.65	Level 1
		6	Oil/grease	0.65	
		7	Wrong colour	0.65	
GENERAL SAFETY AN PERFORMANCE REQUIREMENTS	re	Special characteristic : Product characteristic which can affect safety or compliance with regulations, fit, function, performance or subsequent processing of product. None identified.			
			ribes the properties of product above indicated. This dation, drawing references, defect specification, biologica		

REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (NAME/FUNCTION/SIGNATURE)	APPROVED BY: (NAME/FUNCTION/SIGNATURE)
17/08/2021	1	Initial release.	Kinga Gawdzik – Engineering Support Technician	Andrew Pearce – Quality Manager



Product P/N	2802/30	
	2002/30	
		Mod. 984A
Description	Spiroguard adaptor	Rev. 06
Desemption	ophoguaia adaptor	

CUSTOMER APPROVAL:				
We accept this material specification as a part of the agreed terms of delivery.				
	Company Name:			
Approved by:				
	NAME/FUNCTION			
	SIGNATURE			
	DATE			
	COMPANY STAMP			
Please send bac	this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.			