

# PRODUCT SPECIFICATION

Product P/N	2800/21	Mod. 984A Rev. 06
Description	Spiroguard Integral Mouthpiece	

**2800/21**

## **Spiroguard Integral Mouthpiece**



<b>PRODUCT DESCRIPTION</b>	<p>Inlet/Outlet Connectors:</p> <ul style="list-style-type: none"> <li>- OD34mm ID29.25mm - Machine Side;</li> <li>- Integral Mouthpiece - Patient Side.</li> </ul> <p>Approx. dimensions: 97mm diameter x 93mm height. Weight: 37gm (approx.). Bidirectional Filter.</p>
<b>MANUFACTURER NAME</b>	<p><b>GVS Filter Technology UK</b> NFC House Vickers Industrial Estate Mellishaw Lane, Morecambe Lancashire LA3 3EN - United Kingdom</p> <p><b>Information</b> Tel. +44 (0) 1524 847600 e-mail: gvsuk@gvs.com</p>
<b>INTENDED USE / APPLICATION</b>	<p>Spirometer pre filters are used in pulmonary function tests for trapping bacteria and viruses and other particulates, reducing the risk of cross-contamination between the patient and the machine. Filters can be supplied with accessories, such as nose clip, mouthpiece and bite grip.</p>
<b>CLASS OF THE PRODUCT</b>	<p>Disposable medical device - Class IIa</p> <p>Rule 2 Annex IX 93/42 / EEC</p> <p>Rule 5 Annex VIII MDR 2017/745</p>
<b>MATERIALS</b>	<p><b>Filter media: <i>Electrostatic Blended Synthetic Fiber</i></b> <b>Frame/Housing Polymer: <i>White Polypropylene Homopolymer (PP)</i></b> <b>Colour: <i>White</i>.</b></p> <p><b>Regulatory Documentation Required:</b></p> <ul style="list-style-type: none"> <li>- Biocompatibility according ISO 10993-1</li> <li>- ROHS</li> <li>- BSE/TSE</li> </ul>

# PRODUCT SPECIFICATION

Product P/N	2800/21	Mod. 984A
Description	Spiroguard Integral Mouthpiece	Rev. 06

	<ul style="list-style-type: none"> <li>- DEHP plasticizer Free and latex free</li> <li>- Aging</li> <li>- REACH</li> <li>- Conflict minerals</li> </ul>
PRODUCT CHARACTERISTICS	<p><b>Appearance/Visual</b> As shown on drawing.</p> <p><b>Physical/Mechanical</b> Approx. dimensions: <b>97mm diameter x 93mm height.</b> Weight: <b>37gm (approx.).</b> Interfaces (ex: Input / Output connectors): <b>OD 34mm ID 29.25mm - Machine Side; Integral Mouthpiece - Patient Side.</b></p> <p>Operating temperature Range: <b>N/A</b> Storage temperature Range: <b>5 °C to 40 °C</b> <b>Bidirectional Filter, Male connector – Patient Side.</b></p> <p><b>Biological</b> Pyrogenicity: <b>&lt;0.3 EU/ml</b> <b>Biocompatibility to ISO10993</b> Category – <b>Surface device</b> Contact – <b>Oral cavity</b> Contact Duration - <b>&lt;24hrs</b></p> <p><b>Functional</b> Air Flow Rate: <b>30l/min, 60l/min, 90l/min.</b></p> <p>Filtration Efficiency: <b>Filter Efficiency @ 30L/min using TSI 8130: Min. 99%</b> (REP: 1433/17 with factor of safety applied to Min.)</p> <p>Pressure Drop: <b>Flow Resistance @ 30l/min in accordance with EN ISO 9360-1: Max.35.2Pa</b> <b>Flow Resistance @ 60l/min in accordance with EN ISO 9360-1: Max.75.9Pa</b> <b>Flow Resistance @ 90l/min in accordance with EN ISO 9360-1: Max.127.6Pa</b> (REP:1251/17 with 10% of safety margin added to Max.)</p> <p>Internal Volume: <b>77ml (approx.)</b></p> <p>Operating Lifetime: <b>Refer to Instructions for Use.</b></p> <p>Shelf Lifetime: <b>5 years from the date of manufacture.</b></p> <p><b>Bacterial Filtration Efficiency in accordance with ASTM F2101-07: Min. 99.999%</b> (<i>Staphylococcus aureus</i> @ 30L /minute) REP: EXT607770.</p> <p><b>Viral Filtration Efficiency in accordance with ASTM F2101-07: Min. 99.999%</b> (<i>Bacteriophage</i> @ 30L/ minute) REP: EXT620332.</p> <p><b>Cleanliness</b> Device assembled within Class 8 Cleanroom.</p>


# PRODUCT SPECIFICATION

Product P/N	2800/21	Mod. 984A Rev. 06
Description	Spiroguard Integral Mouthpiece	

	<b>Testing</b> <i>Torque test @ 10Nm. (REP: 1347/17)</i>
<b>INSTRUCTIONS / WARNINGS</b>	<i>Multi-language IFU available.</i>
<b>PRODUCT SHELF LIFE</b>	<i>5 years from the date of manufacture.</i> <i>Expiration date and date of manufacture are detailed on the product labelling.</i>
<b>STERILIZATION</b>	<i>N/A</i>
<b>APPLICABLE STANDARDS AND REGULATIONS</b>	<p><b>Product Certification required:</b></p> <ul style="list-style-type: none"> <li>- CE mark</li> <li>- FDA</li> </ul> <p><b>Applicable Standards and Technical Regulations:</b></p> <p><i>Biological evaluation of Medical Devices - Part 1: Evaluation and Testing - ISO 10993-1.</i></p> <p><i>Respiratory protective devices - Method for test - Part 7: Determination of particle filter penetration - BS EN 13274-7.</i></p> <p><i>Medical devices- Application of risk management to medical devices - BS EN ISO 14971.</i></p> <p><i>Medical devices – symbols to be used with medical device labels, labelling and information to be supplied - Part1: General requirements - ISO 15223-1.</i></p>
<b>PACKAGING AND LABELING</b>	<p><i>Number of pcs per bag is determined by the sales order.</i></p> <p><i>The first barcode label is applied to the outside of the bags.</i></p> <p><i>The second barcode label is applied onto the outside of the box.</i></p> <p><i>Each bag is labelled with the following traceability information:</i></p> <ul style="list-style-type: none"> <li>✓ Quantity</li> <li>✓ Product description</li> <li>✓ Product Date</li> <li>✓ Lot Number (OL and 5-digit batch number to trace back to raw materials used)</li> <li>✓ Operator Code</li> </ul> <p><i>Different lots in one box are separately closed and separately labelled.</i></p> <p><i>Bulk products will be packed in double PE bags.</i></p>
<b>CERTIFICATE OF COMPLIANCE</b>	<p><i>With each shipment, GVS UK Customer Service will send the CoFC to the Customer, based on the lot numbers and date of manufacture.</i></p> <p><i>Conformity declaration is printed on every invoice and Certificate is according to UNI EN 10204 type 2.1.</i></p> <p><i>The Quality management system is in compliance with ISO 9001, ISO 13485.</i></p>
<b>DRAWING</b>	<i>The attached drawing is part of this product specification and must not be duplicated or made accessible to a third party without written permission from GVS Filter Technology UK Ltd.</i>

# PRODUCT SPECIFICATION

Product P/N	2800/21	Mod. 984A
Description	Spiroguard Integral Mouthpiece	Rev. 06

	<p>Approximate dimensions for reference only</p> 																															
ACCEPTABLE QUALITY LEVEL	AQL: 0.65 with sampling Plan: ISO2859.																															
VISUAL REQUIREMENTS	<p><b>Visual acceptance requirements apply when inspected under below conditions:</b></p> <p>Magnification: <i>Unaided eye at a distance of approximately 35-40cm.</i>  Light type: <i>Lighting level must be reasonable for visual detection.</i>  Timings: <i>Maximum inspection period per item is 25 seconds.</i>  <i>For detailed defect list, refer to product control plan.</i></p> <table border="1"> <thead> <tr> <th colspan="2">Acceptance Requirement</th><th>AQL</th><th>Sampling Plan</th></tr> </thead> <tbody> <tr> <td>1</td><td>Black particle contamination</td><td>0.65</td><td rowspan="8">ISO 2859 Part 1 General Inspection Level 1</td></tr> <tr> <td>2</td><td>Damaged/broken item</td><td>0.65</td></tr> <tr> <td>3</td><td>Blocked connector/luer</td><td>0.65</td></tr> <tr> <td>4</td><td>Weld marks</td><td>0.65</td></tr> <tr> <td>5</td><td>Short fill moulding</td><td>0.65</td></tr> <tr> <td>6</td><td>Rough surface or edges</td><td>0.65</td></tr> <tr> <td>7</td><td>Pronounced injection gate</td><td>0.65</td></tr> <tr> <td>8</td><td>Deformation/distortion</td><td>0.65</td></tr> </tbody> </table>			Acceptance Requirement		AQL	Sampling Plan	1	Black particle contamination	0.65	ISO 2859 Part 1 General Inspection Level 1	2	Damaged/broken item	0.65	3	Blocked connector/luer	0.65	4	Weld marks	0.65	5	Short fill moulding	0.65	6	Rough surface or edges	0.65	7	Pronounced injection gate	0.65	8	Deformation/distortion	0.65
Acceptance Requirement		AQL	Sampling Plan																													
1	Black particle contamination	0.65	ISO 2859 Part 1 General Inspection Level 1																													
2	Damaged/broken item	0.65																														
3	Blocked connector/luer	0.65																														
4	Weld marks	0.65																														
5	Short fill moulding	0.65																														
6	Rough surface or edges	0.65																														
7	Pronounced injection gate	0.65																														
8	Deformation/distortion	0.65																														

# PRODUCT SPECIFICATION

Product P/N	2800/21	Mod. 984A Rev. 06
Description	Spiroguard Integral Mouthpiece	

	9	Crack	0.65		
	10	Oil/grease	0.65		
	11	Wrong colour	0.65		
	12	Weld fault	0.65		
GENERAL SAFETY AND PERFORMANCE REQUIREMENTS	<b>Special characteristic:</b> <i>Product characteristic which can affect safety or compliance with regulations, fit, function, performance or subsequent processing of product.</i>				
	<b>Special Characteristic # 01:</b>				
	<i>Flow Resistance @ 30L/min in accordance with EN ISO 9360-1</i>				
	<i>Flow Resistance @ 60L/min in accordance with EN ISO 9360-1</i>				
	<i>Flow Resistance @ 90L/min in accordance with EN ISO 9360-1</i>				
	<b>Special Characteristic # 02:</b> <i>Filter Efficiency @ 30L/min using TSI 8130 in accordance with EN 13274-7.</i>				
<b>Special Characteristic # 03:</b> <i>Bacterial Filtration Efficiency in accordance with ASTM F2101-07</i>					
<i>Viral Filtration Efficiency in accordance with ASTM F2101-07</i>					
This material specification describes the properties of product above indicated. This document contains general requirements, material description, drawing references, defect specification, biological material requirements.					

## REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (NAME/FUNCTION/SIGNATURE)	APPROVED BY: (NAME/FUNCTION/SIGNATURE)
10/08/2021	3	IFU languages section and internal volume amended.	Kinga Gawdzik – Engineering Support Technician 	Andrew Pearce – Quality Manager 

# PRODUCT SPECIFICATION

<b>Product P/N</b>	2800/21	<b>Mod. 984A Rev. 06</b>
<b>Description</b>	Spiroguard Integral Mouthpiece	

## CUSTOMER APPROVAL:

We accept this material specification as a part of the agreed terms of delivery.

Company Name:

Approved by:

NAME/FUNCTION

SIGNATURE

DATE

COMPANY STAMP

*Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.*