# PRODUCT SPECIFICATION

| | | |
|---|---|---|
| **Product P/N** | **2200/902** | **Mod. 984A Rev. 06** |
| **Description** | **90mm Suction Filter** | |

## *2200/902*

## *90mm Suction Filter*

| | |
|---|---|
| **PRODUCT DESCRIPTION** | Inlet/Outlet Connectors: 11mm to 15mm hose barbed both sides.<br>Approx. dimensions: 92mm diameter x 61mm height.<br>Flow direction indicated by “IN” marking on filter face.<br>Weight: 47gm (approx.).<br>Unidirectional filter. |
| **MANUFACTURER NAME** | **GVS Filter Technology UK**<br>NFC House<br>Vickers Industrial Estate<br>Mellishaw Lane, Morecambe<br>Lancashire LA3 3EN - United Kingdom<br><br>**Information**<br>Tel. +44 (0) 1524 847600<br>e-mail: gvsuk@gvs.com |
| **INTENDED USE / APPLICATION** | The hydrophobic filter protects surgical suction vacuum pump equipment from contamination. |
| **CLASS OF THE PRODUCT** | Disposable medical device - Class IIa<br>Rule 2 Annex IX 93/42 / EEC<br>Rule 2 Annex VIII MDR 2017/745 |
| **MATERIALS** | **Filter media:** ***Hydrophobic PTFE 1.0µm***<br><br>**Frame/Housing Polymer:** ***Clear Styrene-Butadiene Copolymer (SBC)***<br><br>**Colour:** ***Transparent Clear***<br><br>**Regulatory Documentation Required:**<br>- Biocompatibility according ISO 10993-1<br>- ROHS<br>- BSE/TSE<br>- DEHP plasticizer Free and latex free<br>- Aging<br>- REACH<br>- Conflict minerals |

# PRODUCT SPECIFICATION

| | | |
|---|---|---|
| **Product P/N** | **2200/902** | **Mod. 984A Rev. 06** |
| **Description** | **90mm Suction Filter** | |

| | |
|---|---|
| **PRODUCT CHARACTERISTICS** | **Appearance/Visual**<br>As shown on drawing.<br><br>**Physical/Mechanical**<br>*Approx. dimensions:* ***92mm diameter x 61mm height.***<br>*Weight:* ***47gm (approx.).***<br>Interfaces (ex: Input / Output connectors): ***11mm to 15mm hose barbed both sides.***<br>Operating temperature Range: ***N/A***<br>Storage temperature Range: ***5 °C to 40 °C***<br>***Unidirectional Filter.***<br>***Flow direction indicated by "IN" marking on filter face.***<br><br>**Biological**<br>Pyrogenicity: ***<0.3 EU/ml***<br>***Biocompatibility to ISO10993***<br>*Category –* ***Surface device***<br>*Contact –* ***Skin***<br>*Contact Duration -* ***<24hrs***<br><br>**Functional**<br>Air Flow Rate: ***Min. 54.4 l/min @ 1PSI*** (REP: 2184/20 with 20% Factor of Safety applied to Min.)<br><br>Water Break Through Pressure: ***0.8bar (600mmHG) Vacuum for 20s*** (REP: 2186/20)<br><br>Filtration Efficiency:<br>*Filter Efficiency @ 30L/min using DEHS@0.3µm****: Min.99.995%***<br>(Ref. 2200/911 REP: 2183/20 with Factor of Safety applied to Min.)<br><br>Pressure Drop: ***N/A***<br>Internal Volume: ***N/A***<br>Operating Lifetime: ***Refer to Instructions for Use.***<br>Shelf Lifetime: ***5 years from the date of manufacture.***<br><br>*Bacterial Filtration Efficiency in accordance with ASTM F2101-07:* ***Min. 99.999%***<br>*(Staphylococcus aureus @ 30L /minute)* Ref.2200/02 REP: EXT846022-S01<br><br>*Viral Filtration Efficiency in accordance with ASTM F2101-07****: Min. 99.999%***<br>*(Bacteriophage @ 30L/ minute)* Ref.2200/02 REP: EXT846021-S01<br><br>**Cleanliness**<br>Device assembled within Class 8 Cleanroom.<br><br>**Testing**<br>Burst test: ***Min. 16.7psi*** (REP: 2185/20 with 20% Factor of Safety applied to Min.) |
| **INSTRUCTIONS / WARNINGS** | *Multi-language IFU available.* |
| **PRODUCT SHELF LIFE** | *5 years from the date of manufacture.*<br><br>*Expiration date and date of manufacture are detailed on the product labelling.* |

# PRODUCT SPECIFICATION

| | | |
|---|---|---|
| **Product P/N** | **2200/902** | **Mod. 984A Rev. 06** |
| **Description** | **90mm Suction Filter** | |

| | |
|---|---|
| **STERILIZATION** | *N/A* |
| **APPLICABLE STANDARDS AND REGULATIONS** | **Product Certification required:**<br>- CE mark<br>- FDA<br><br>**Applicable Standards and Technical Regulations:**<br>*Biological evaluation of Medical Devices - Part 1: Evaluation and Testing - ISO 10993-1.*<br><br>*Medical devices- Application of risk management to medical devices - BS EN ISO 14971.*<br><br>*Medical devices – symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements - ISO 15223-1.*<br><br>*High Efficiency Air Filters – BS EN 1822.* |
| **PACKAGING AND LABELING** | *Number of pcs per bag is determined by the sales order.*<br>*The first barcode label is applied to the outside of the bags.*<br>*The second barcode label is applied onto the outside of the box.*<br>*Each bag is labelled with the following traceability information:*<br>✓ *Quantity*<br>✓ *Product description*<br>✓ *Product Date*<br>✓ *Lot Number (OL and 5-digit batch number to trace back to raw materials used)*<br>✓ *Operator Code*<br><br>*Different lots in one box are separately closed and separately labelled.*<br>*Bulk products will be packaged in double PE bags* |
| **CERTIFICATE OF COMPLIANCE** | *With each shipment, GVS UK Customer Service will send the CofC to the Customer, based on the lot numbers and date of manufacture.*<br>*Conformity declaration is printed on every invoice and Certificate is according to UNI EN 10204 type 2.1.*<br><br>*The Quality management system is in compliance with ISO 9001, ISO 13485.* |
| **DRAWING** | *The attached drawing is part of this product specification and must not be duplicated or made accessible to a third party without written permission from GVS Filter Technology UK Ltd.* |

# PRODUCT SPECIFICATION

| | | |
|---|---|---|
| **Product P/N** | **2200/902** | **Mod. 984A Rev. 06** |
| **Description** | **90mm Suction Filter** | |







Approximate dimensions for reference only

**ACCEPTABLE QUALITY LEVEL**

AQL: with sampling Plan: 0.65 AQL to ISO 2859-1

**VISUAL REQUIREMENTS**

**Visual acceptance requirements apply when inspected under below conditions:**

Magnification: *Unaided eye at a distance of approximately 35-40cm.*
Light type: *Lighting level must be reasonable for visual detection.*
Timings: *Maximum inspection period per item is 25 seconds.*
*For detailed defect list, refer to product control plan.*

| | Acceptance Requirement | AQL | Sampling Plan |
|---|---|---|---|
| 1 | Black particle contamination | 0.65 | ISO 2859 Part 1 General Inspection Level 1 |
| 2 | Damaged/broken item | 0.65 | |
| 3 | Blocked connector/luer | 0.65 | |

| | | | | |
|---|---|---|---|---|
| | 4 | Short fill moulding | 0.65 | |
| | 5 | Rough surface or edges | 0.65 | |
| | 6 | Pronounced injection gate | 0.65 | |
| | 7 | Deformation/distortion | 0.65 | |
| | 8 | Crack | 0.65 | |
| | 9 | Oil/grease | 0.65 | |
| | 10 | Wrong colour | 0.65 | |
| | 11 | Weld fault | 0.65 | |
| | 12 | Weld marks | 0.65 | |

| | |
|---|---|
| **GENERAL SAFETY AND PERFORMANCE REQUIREMENTS** | **Special characteristic**: *Product characteristic which can affect safety or compliance with regulations, fit, function, performance or subsequent processing of product.*<br><br>**Special Characteristic # 01:** *Air Flow rate at 1psi.*<br>**Special Characteristic # 02:** *Filter Efficiency @ 30L/min using DEHS @ 0.3µm in accordance with High Efficiency Air Filters – BS EN 1822.*<br>**Special Characteristic # 03:** *Bacterial Filtration Efficiency in accordance with ASTM F2101-07*<br>*Viral Filtration Efficiency in accordance with ASTM F2101-07*<br>**Special Characteristic # 04:** *Burst Pressure.* |

**This material specification describes the properties of product above indicated. This document contains general requirements, material description, drawing references, defect specification, biological material requirements.**

## REVISIONS AND APPROVALS:

| DATE | REV. | REASON FOR CHANGE | ISSUED AND CONTROLLED BY: (NAME/FUNCTION/SIGNATURE) | APPROVED BY: (NAME/FUNCTION/SIGNATURE) |
|---|---|---|---|---|
| 07/09/2021 | 2 | IFU section updated on Mod.984. | Kinga Gawdzik - Engineering Support Technician  | Andrew Pearce – Quality Manager  |

# PRODUCT SPECIFICATION

| | | |
|---|---|---|
| **Product P/N** | **2200/902** | **Mod. 984A Rev. 06** |
| **Description** | **90mm Suction Filter** | |

## CUSTOMER APPROVAL:

**We accept this material specification as a part of the agreed terms of delivery.**

**Company Name:**

**Approved by:**

**NAME/FUNCTION**

**SIGNATURE**

**DATE**

**COMPANY STAMP**

*Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.*