

<b>Product PN</b>	FJ25BNPRC002AD01, FJ25BNPRC004AD01	<b>Mod.</b> 984
<b>Description</b>	25 mm ABLUO Syringe Filters w/Regenerated Cellulose (RC)	<b>Rev.</b> 02





## 25 mm ABLUO, Regenerated Cellulose (RC)



<b>PRODUCT DESCRIPTION</b>	Non-sterile 25 mm syringe filter made of Polypropylene housing, assembled with various pore sizes of Regenerated Cellulose (RC) Membrane						
	Membrane Material	Pore Size (um)	End Fitting	Color	Housing Material	Packaging	Product Code
	Regenerated Cellulose (RC)	0.20	FLL/MLS	Transparent	Polypropylene	500/pk	FJ25BNPRC002AD01
	Regenerated Cellulose (RC)	0.45	FLL/MLS	Transparent	Polypropylene	500/pk	FJ25BNPRC004AD01
<b>MANUFACTURER NAME</b>	<b>GVS North America</b> <b>63 Community Drive</b> <b>Sanford, Me 04073</b> Phone: +1.866.736.1250 eMail: <a href="mailto:CustomerCareGVSLs@gvs.com">CustomerCareGVSLs@gvs.com</a> - Website: <a href="http://www.gvs.com">www.gvs.com</a>						
<b>INTENDED USE / APPLICATION</b>	<b>Applications</b> <ul style="list-style-type: none"> <li>Analytical sample preparation</li> <li>Biological fluids</li> <li>Buffer solutions</li> <li>Sterile filtering of tissue culture media</li> <li>Protein aqueous solutions</li> </ul>						
<b>MATERIALS</b>	<b>Filter media:</b> Regenerated Cellulose (RC) <b>Frame/Housing Polymer:</b> Polypropylene <b>Color:</b> Transparent <b>Other insert(s):</b> N/A <b>Regulatory Documentation Required:</b> <ul style="list-style-type: none"> <li><input checked="" type="checkbox"/> Biocompatibility according ISO 10993-1</li> <li><input type="checkbox"/> IMDS</li> <li><input type="checkbox"/> DEHP plasticizer Free and latex free</li> <li><input checked="" type="checkbox"/> Rohs, Directive 2002/32/CE</li> <li><input type="checkbox"/> Aging</li> <li><input checked="" type="checkbox"/> BSE/TSE, directive 2003/32/CE</li> <li><input checked="" type="checkbox"/> 1907/2006/CE (hazardous substances regulation)</li> <li><input type="checkbox"/> Dir. 67/548/CE and Reg. 1272/2008/CE (medical sector dangerous substances)</li> <li><input checked="" type="checkbox"/> Conflict minerals</li> </ul>						

<b>Product PN</b>	FJ25BNPRC002AD01, FJ25BNPRC004AD01	<b>Mod.</b> 984
<b>Description</b>	<b>25 mm ABLUO Syringe Filters w/Regenerated Cellulose (RC)</b>	<b>Rev.</b> 02

<b>PRODUCT CHARACTERISTIC</b>	<p>Membrane Materials: Regenerated Cellulose          Membrane Diameter: 25 mm          Effective Filtration Area: 4.6 cm<sup>2</sup>          Housing Diameter: 33 mm          Housing Materials: Polypropylene          Inlet / Outlet: FLL / MLS          Holdup Volume: &lt;100 microliter          Maximum Operating Temperature: PP Abluo - 90°C/194°F,          Maximum Operating Pressure: 80 psi          Sterile: No</p>
<b>PRODUCT SHELF LIFE</b>	When stored under normal storage conditions, this product should be stable for 5 years
<b>STERILIZATION</b>	<input type="checkbox"/> EtO <input type="checkbox"/> Gamma <input type="checkbox"/> Beta <input type="checkbox"/> Steam <input type="checkbox"/> e-beam <input checked="" type="checkbox"/> Not Required
<b>COMPLIANCE</b>	The Quality management system is in compliance with ISO 9001:2008, ISO/TS 16949:2009
<b>DRAWING</b>	<div style="display: flex; justify-content: space-around; align-items: flex-start;"> <div style="text-align: center;">  <p>1. Male Luer Slip Outlet (ISO 594)            2. RC Membrane            3. Female Luer Lock Inlet (ISO 594)</p> </div> <div style="text-align: center;">  <p>SECTION A-A</p> </div> </div> <p style="text-align: right; margin-top: 10px;">All Dimensions are in mm, for ref. only</p>

# PRODUCT SPECIFICATION



<b>Product PN</b>	FJ25BNPRC002AD01, FJ25BNPRC004AD01	<b>Mod.</b> 984
<b>Description</b>	25 mm ABLUO Syringe Filters w/Regenerated Cellulose (RC)	<b>Rev.</b> 02

<b>VISUAL REQUIREMENTS</b>	<i>Visual acceptance requirements apply when inspected under below conditions:</i>			
	<b>Magnification: None</b>			
	<b>Light type: Standard</b>			
	Acceptance Requirement		Sampling Plan	
	1	Contamination	None	100%
	2	Damaged Luer Fitting	None	100%
	3	Missing Membrane	None	100%
	4	Incomplete Membrane	None	100%
	5	Membrane Displacement	None	100%
	6	Membrane Protruding Out of Part	None	100%
	7	Scuffed Surface	Total length of scuff exceed 2 ribs	100%
	8	Cracked Housings	None	100%
	9	Weld Flash	None	100%
10	Burn Outside of the Stake Ring	None	100%	
11	Embedded Particles	< 0.8 mm <sup>2</sup> (Maximum 3 particles)	100%	
12	Mold Flash	< 0.2 mm	100%	

<b>PERFORMANCE REQUIREMENTS</b>	Acceptance Requirement		Sampling Plan	
	Pore size	0.20 um	0.45 um	AQL 0.1 Special inspection level S3
	Pressure	≥ 80 PSI 10 Seconds	≥ 80 PSI 10 Seconds	ANSI/ASQ Standard Z1.4 - 2008
	Min. Bubble point (psi)	45	32	AQL 0.1 Special inspection level S3

This material specification describes the properties of product above indicated.  
 This document contains general requirements, material description, drawing references, defect specification, biological material requirements.

## REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (name /function and signature)	APPROVED BY: (name /function and signature)
3/31/17	0	Initial Release	Joe DeSisto, Director, Process Engineering 	Kevin Wrigley, Director, Quality 



# PRODUCT SPECIFICATION

<b>Product PN</b>	FJ25BNPRC002AD01, FJ25BNPRC004AD01	<b>Mod.</b> 984
<b>Description</b>	25 mm ABLUO Syringe Filters w/Regenerated Cellulose (RC)	<b>Rev.</b> 02

## *Customer Approval:*

We accept this material specification as a part of the agreed terms of delivery

**Company name** \_\_\_\_\_

**Approved by:** \_\_\_\_\_  
(Name, Function) (Signature)

**Date** \_\_\_\_\_  
(Company stamp)

*Please send back this document signed for approval. If we will not receive this specification signed , we consider the first order placed as implicit approval.*