

Product P/N	A620/60/61	
		Mod. 984A
Description	Attachment	Rev. 06



Attachment



PRODUCT	Inlet/Outlet Connectors:
DESCRIPTION	- connector ISO 22mm Male / 15mm Female - 15mm Male,
	- connector ISO 22mm Male/ 22mm Barbed,
	- Collapsible tube with connectors 15.25mm and 22.25mm.
	Approx. dimensions: 179mm length.
	Weight: 20g (approx.).
MANUFACTURER	GVS Filter Technology UK
NAME	NFC House
	Vickers Industrial Estate
	Mellishaw Lane, Morecambe
	Lancashire LA3 3EN - United Kingdom
	Information
	Tel. +44 (0) 1524 847600
	e-mail: gvsuk@gvs.com
INTENDED USE /	Catheter mounts are used in conjunction with filters in ventilator patient circuits. For performance
APPLICATION	data, refer to the Instruction For Use for associated filter.
CLASS OF THE	Disposable medical device - Class Ila
PRODUCT	Rule 2 Annex IX 93/42 / EEC
	Rule 2 Annex VIII MDR 2017/745
MATERIALS	Frame/Housing Polymer: Tube: Transparent clear Polypropylene (PP)
	Connector: Clear Styrene - Butadiene Copolymer (SBC)
	Colour: Transparent Clear.
	Regulatory Documentation Required:
	- Biocompatibility according ISO 10993-1
	- ROHS
	- BSE/TSE
	 DEHP plasticizer Free and latex free
	- Aging
	- REACH
	- Conflict minerals
DRODUCT	
PRODUCT CHARACTERISTICS	Appearance/Visual
CHARACTERISTICS	As shown on drawing.



PRODUCT SPECIFICATION

Product P/N	A620/60/61	Mod. 984A
Description	Attachment	Rev. 06

	Physical/Mechanical Approx. dimensions: 179mm length. Weight: 20g (approx.) Interfaces (ex: Input / Output connectors): - Connector ISO 22mm Male / 15mm Female - 15mm Male; - Connector ISO 22mm Male/ 22mm Barbed; - Collapsible tube with connectors 15.25mm and 22.25mm, 128mm length (when un- collapsed).
	Operating temperature Range: N/A Storage temperature Range: 5 °C to 40 °C
	Biological Biocompatibility to ISO10993 Category – Surface device Contact – Skin Contact Duration - <24hrs
	Functional Air Flow Rate: N/A Filtration Efficiency: N/A Pressure Drop: N/A Internal Volume: 49ml (approx.) Operating Lifetime: Refer to Instructions for Use. Shelf Lifetime: 5 years from the date of manufacture.
	Cleanliness Device assembled within Class 8 Cleanroom.
	Testing N/A
INSTRUCTIONS / WARNINGS	Multi-language IFU available.
PRODUCT SHELF LIFE	5 years from the date of manufacture.
STERILIZATION	Expiration date and date of manufacture are detailed on the product labelling. Sterile version of product available (Ethylene oxide - Max 55°C).
APPLICABLE STANDARDS AND REGULATIONS	Product Certification required: - CE mark - FDA
	Applicable Standards and Technical Regulations: Biological evaluation of Medical Devices - Part 1: Evaluation and Testing - ISO 10993-1.
	Medical devices- Application of risk management to medical devices - BS EN ISO 14971.
	Medical devices – symbols to be used with medical device labels, labelling and information to be supplied - Part1: General requirements - ISO 15223-1.



Product P/N	A620/60/61	
		Mod. 984A
Description	Attachment	Rev. 06

	Sterilization of health care products – Ethylene oxide sterilization – ISO 11135-1.		
	Sterilization of medical devices – Microbiological Methods – Part 1: Estimation of population of Microorganisms on products – ISO 11737-1.		
PACKAGING AND LABELING	Number of pcs per bag is determined by the sales order. The first barcode label is applied to the outside of the bags. The second barcode label is applied onto the outside of the box. Each bag is labelled with the following traceability information: ✓ Quantity ✓ Product description ✓ Product Date ✓ Lot Number (OL and 5-digit batch number to trace back to raw materials used) ✓ Operator Code Different lots in one box are separately closed and separately labelled. Bulk products will be packed in double PE bags.		
CERTIFICATE OF COMPLIANCE	With each shipment, GVS UK Customer Service will send the CofC to the Customer, based on the lot numbers and date of manufacture. Conformity declaration is printed on every invoice and Certificate is according to UNI EN 10204 type 2.1. The Quality management system is in compliance with ISO 13485.		
DRAWING	The attached drawing is part of this product specification and must not be duplicated or made accessible to a third party without written permission from GVS Filter Technology UK Ltd. Approximate dimensions for reference only		
ACCEPTABLE QUALITY LEVEL	AQL: 0.65 with sampling Plan: ISO2859.		
VISUAL REQUIREMENTS	Visual acceptance requirements apply when inspected under below conditions: Magnification: Unaided eye at a distance of approximately 35-40cm. Light type: Lighting level must be reasonable for visual detection. Timings: Maximum inspection period per item is 25 seconds. For detailed defect list, refer to product control plan.		



Product P/N	A620/60/61	Mod. 984A
Description	Attachment	Rev. 06

		Acceptance Requirement	AQL	Sampling Plan
	:	Black particle contamination	0.65	
		Damaged/broken item	0.65	
	:	Blocked connector/luer	0.65	
	4	Short fill moulding	0.65	
	5	Rough surface or edges	0.65	ISO 2859 Part 1 General Inspection
	(Pronounced injection gate	0.65	Level 1
	-	Deformation/distortion	0.65	
	5	Crack	0.65	
	9	Oil/grease	0.65	
	1	0 Wrong colour	0.65	
GENERAL SAFETY AND PERFORMANCE REQUIREMENTS	Special characteristic : Product characteristic which can affect safety or compliance with regulations, fit, function, performance or subsequent processing of product.			

This material specification describes the properties of product above indicated. This document contains general requirements, material description, drawing references, defect specification, biological material requirements.

REVISIONS AND APPROVALS:

				•
DATE	REV.	REASON	ISSUED AND CONTROLLED BY:	APPROVED BY:
		FOR	(NAME/FUNCTION/SIGNATURE)	(NAME/FUNCTION/SIGNATURE)
		CHANGE		
02/08/2021	2	Contact and sterile sections amended.	Kinga Gawdzik – Engineering Support Technician	Andrew Pearce – Quality Manager
1				



Product P/N	A620/60/61	Mod. 984A
Description	Attachment	Rev. 06

CUSTOMER APPROVAL:			
We accept this	We accept this material specification as a part of the agreed terms of delivery.		
Company Nam	ne:		
Approved by:	NAME/FUNCTION		
	SIGNATURE		
	DATE		
	COMPANY STAMP		

Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.