

MEMBRANE SPECIFICATION

Product PN **M09G0300HxxxxE ***

Mod. 984 d

Description **HYDROPHOBIC, HAEMOREPELLENT 3.0 µm FORTEX MEMBRANE**

Rev. 04

HYDROPHOBIC AND HAEMOREPELLENT 3.0 µm PVDF FORTEX MEMBRANES



PRODUCT DESCRIPTION	PVDF hydrophobic and haemorepellent 3.0 µm membranes cast on a nonwoven polyester support. Two-sided. Provided in rolls of different widths.						
APPLICATIONS	The product is ideal for venting in medical applications.						
MATERIALS COMPONENTS	<table> <tr> <td>Polymer:</td><td>PVDF homopolymer</td></tr> <tr> <td>Support:</td><td>Polyester non-woven</td></tr> <tr> <td>Others</td><td>Haemorepellent treatment</td></tr> </table>	Polymer:	PVDF homopolymer	Support:	Polyester non-woven	Others	Haemorepellent treatment
Polymer:	PVDF homopolymer						
Support:	Polyester non-woven						
Others	Haemorepellent treatment						

PRODUCT CHARACTERISTICS

Appearance/Visual:
White porous ribbons

Dimensional:

Standard Product code	Width*	Thickness	Max roll diameter	Standard length	Min length roll	Max length roll	Internal core diameter
M09G0300H0254E	25.4 mm	150 – 200 µm	220 mm	125 m	90 m	180 m	76 ± 0.8 mm
M09G0300H0340E	34.0 mm						
M09G0300H0381E	38.1 mm						
M09G0300H0787E	78.7 mm						
M09G0300H1524E	152.4 mm						
M09G0300H2540E	254.0 mm						
M09G0300H3048E	304.8 mm						
M09G0300H3302E	330.2 mm						
M09G0030H4500E	450.0 mm						

For other available widths, please contact the sales representative of your country.
The following product specification is valid also for other part numbers M09G0300xxxxE that do not appear in the table above.

* For width ≤ 15.5 mm : Tolerance ± 0.2 mm
For width > 15.5 mm : Tolerance ± 0.8 mm

MEMBRANE SPECIFICATION

Product PN **M09G0300HxxxxE ***

Mod. 984 d

Description **HYDROPHOBIC, HAEMOREPELLENT 3.0 µm FORTEX MEMBRANE**

Rev. 04

Biological:

Test	Standards/ procedures of ref	Result
Bacteria filtration efficiency test (BFE) using <i>Staphylococcus aureus</i>	ASTM F2101	99%
Chemical toxicity and solvent detection	UNI EN ISO 1135-4 (Annex B) UNI EN ISO 8536-4 (Annex B)	Non toxic
Cytotoxicity-MEM elution in vitro	UNI EN ISO 10993-5 (Annex C)	Non cytotoxic
Intracutaneous reactivity test	UNI EN ISO 10993-10	Non-irritant
Delayed hypersensitivity test (Guinea-pig maximization test - GPMT)	UNI EN ISO 10993-10	0% Sensitization
Acute systemic toxicity	UNI EN ISO 10993-11	Non-toxic/non-irritant
In vitro haemolysis test	UNI EN ISO 10993-4 according to Italian Pharmacopoeia	Non hemolytic
Haemocompatibility	UNI EN ISO 10993-4	Hemocompatible

Regulatory Documentation

Reg. 1907/2006/EC (REACH)
Directive 2011/65/EU (RoHS)
BPA, DEHP and Latex free
Directive 2003/32/EC (TSE-BSE)
UNI EN ISO 10993-1 (Biocompatibility)
Directive 67/546/EC and Reg.1272/2008/EC (Medical sector dangerous substances)

Physical:

Test	Result
Air Flow GVS Internal GVS internal procedure IC303	AF ≥ 42.5 l/min on 1 cm² at 1 bar
Water Break Through GVS internal procedure IC305	≥ 0.15 bar
Haemorepellency GVS internal procedure IC660	Yes
Oleorepellency grade AATCC 118-2002	≥ 3.5
Hydrorepellency grade DuPont Teflon test	≥ 9.5

STERILIZATION COMPATIBILITY

The product is NOT provided in sterile conditions (compatible sterilization method listed below)

Method
<input checked="" type="checkbox"/> ETO
<input checked="" type="checkbox"/> Gamma (25 kGy)
<input checked="" type="checkbox"/> Autoclave (121°C)

PRODUCT SHELF LIFE

5 years

* M09G0300HxxxxE, where xxxx is the width of the roll. For instance xxxx=3302 means 330.2 mm

MEMBRANE SPECIFICATION

Product PN **M09G0300HxxxxE ***

Mod. 984 d

Description **HYDROPHOBIC, HAEMOREPELLENT 3.0 µm FORTEX MEMBRANE**

Rev. 04

STORAGE

Keep the product in a closed and clean environment, away from direct light and excessive temperature and humidity conditions

PACKAGING AND LABELLING

Each roll is individually packed in a thermo-welded PE bag identified with a barcode label.
A second PE bag collects all rolls with the same lot in a cardboard box.
Another barcode label is applied outside the box, with reference to the cumulative quantity.
Each label reports the following traceability information:

- Quantity
- Product description
- Product date and time
- Lot number (OL and 6 digits batch number to trace back to raw materials used)
- Operator code

An image of a barcode label is provided below as an example.



CERTIFICATE OF COMPLIANCE

Conformity declaration is printed on every invoice with a statement according to UNI EN 10204 type 2.1.
The Quality management system in compliance with ISO 9001.

VISUAL REQUIREMENTS

Visual acceptance requirements apply when inspected under below conditions:

Light type: unaided eye + lamp

Distance: 30 – 45 cm

Timings: 5 sec per unit

Acceptance Requirement		AQL	Sampling Plan
1	Embedded contaminations	For widths ≤ 20mm ->Max 2/lm For 20 mm < width ≤ 60mm -> Max 5/lm	First 3 linear meters of rolls sampled according ISO 2859 part. 1 1 st Level
2	Loose contaminations	For 60 mm < width ≤ 130mm ->Max 10/lm For 130 mm < width ≤ 450mm ->Max 20/lm	
3	Grease, stains	0 defects	
4	Splices/cross-tape for each rolls	Max 7	

* M09G0300HxxxxE, where xxxx is the width of the roll. For instance xxxx=3302 means 330.2 mm

MEMBRANE SPECIFICATION

Product PN **M09G0300HxxxxE ***

Mod. 984 d

Description **HYDROPHOBIC, HAEMOREPELLENT 3.0 µm FORTEX MEMBRANE**

Rev. 04

This material specification describes the properties of product above indicated.
This document contains general requirements, material description, drawing references, defect specification, biological material requirements.

REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	PREPARED BY: (name /function and signature)	CONTROLLED BY: (name /function and signature)	APPROVED BY: (name /function and signature)
11/05/2015	02	This specification is extended to all the heights available for the rolls and replaces previous rev.01 for M09G0300H1309E.	Sara Zaccheroni/AQP	Rinaldi Davide/RIM	Luca Querzè/RPT
29/09/2015	03	Added compliance to conflict minerals; new layout	Sara Zaccheroni/AQP	Rinaldi Davide/RIM	Luca Querzè/RPT
13/11/2015	04	Added compliance to ISO 8536-4 for chemical toxicity, FORTEX brand and specs for the new height M09G0300H0127E	Sara Zaccheroni/AQP	Rinaldi Davide/RIM	Luca Querzè/RPT
11/12/2015	05	Added new widths and updated visual and biological requirements	Sara Zaccheroni/AQP	Rinaldi Davide/RIM	Luca Querzè/RPT
01/08/2016	06	Added new widths and updated layout	Sara Zaccheroni/AQP	Rinaldi Davide/ RIM	Luca Querzè/RPT
17/10/2019	07	Modified the widths and updated parameters	Brunella Merante/ R&D <i>Brunella Merante</i>	Annarita Trotta/AQP <i>Annarita Trotta</i>	Luca Querzè/RPT <i>Luca Querzè</i>

Customer Approval:

We accept this material specification as a part of the agreed terms of delivery

Company name _____

Approved by: _____
(Name, Function) (Signature)

Date _____
(Company stamp)

Please send back this document signed for approval. If we will not receive this specification signed, we consider the first order placed as implicit approval.