

<b>Product PN</b> GMD019A00S	<b>Mod. 984e-ext</b>  <b>Rev. 07</b>
<b>Description</b> EPI-Baby IV Filter Neonatal 0.2 µm Luer Lock	



## EPI-Baby IV Filter Neonatal 0.2 µm with Luer Lock

<b>PRODUCT DESCRIPTION</b>	<p>Non-vented filter for infusion and anaesthesia. The device is sterile and can be used in combination with an infusion set or an infusion device (pump or syringe). The connections are a female luer lock with vented cap inlet and a rotating male luer lock with vented cap outlet, compliant with ISO 80369-7. The device is provided with a non-vented EPI-Baby Neonatal filter, with an hydrophilic PES membrane with 0.2 µm pore size for particles retention. The filter size (LxWxH) is 15.3x62x9.6 mm.</p>	
<b>MANUFACTURER NAME</b>	<p><b>GVS SpA</b> Via Roma, 50 40069 Zola Predosa (BO) – ITALY Phone: +39.051.6176311 – Fax: +39.051.6176200 e-mail: <a href="mailto:gvs@gvs.it">gvs@gvs.it</a> – website: <a href="http://www.gvs.com">www.gvs.com</a></p>	
<b>INTENDED USE / APPLICATION</b>	<p>Infusion of Drugs (including chemotherapy), Anesthetics, Nutritional Solutions and Saline Solutions. The filter can be connected via Luer Locks to all commercially available infusion administration sets compliant with ISO 80369-7. The filter can be used in combination with an infusion set or an infusion device (pump or syringe). The filter is intended specifically to eliminate air bubbles and to retain particles and bacteria (0.2 µm). The filter can be used for applications that last up to 96 hours. The device has to be disposed after each therapy. The device should only be supervised and used by qualified healthcare personnel. The filter can't be used for infusion of hyper alimentation liquids, lipids, blood and/or blood derivatives.</p>	
<b>CLASS OF THE PRODUCT</b>	<p>Class IIa – sterile – single use Rule 3 Annex VIII Regulation (UE) 2017/745</p>	
<b>EMDN</b>	<p>A04010101 WITHDRAWAL AND ADMINISTRATION FILTERS (0.2-1.2 MICRONS)</p>	
<b>MATERIALS</b>	<p><b>Filter media:</b> PES <b>Filter Housing:</b> PP, MABS <b>Caps:</b> HDPE</p> <p><b>Regulatory Compliance</b></p> <ul style="list-style-type: none"> <li>▪ Biocompatibility according to ISO 10993-1</li> <li>▪ Rohs directive 2011/65/UE</li> <li>▪ DEHP plasticizer Free</li> <li>▪ Latex free</li> <li>▪ Reach 1907/2006/CE (hazardous substances regulation)</li> <li>▪ Reg. 1272/2008/CE (classification, labelling and packaging of substances and mixtures)</li> </ul>	
<b>PRODUCT CHARACTERISTIC</b>	<p><b>Physical/Mechanical:</b></p> <p><b>Dimensions (LxWxH):</b> 15.3x62x9.6 mm</p> <p><b>Total internal volume of the set:</b> &lt; 0.35 ml</p> <p><b>Weight:</b> 2.75 g</p> <p><b>Input/output connectors:</b> Female luer lock with vented cap inlet and rotating male luer lock with vented cap outlet, compliant with ISO 80369-7</p> <p><b>Operating temperature Range:</b> From 5 °C to 40 °C</p> <p><b>Storage temperature Range:</b> From 0 °C to 40 °C</p>	



# PRODUCT SPECIFICATION



<b>Product PN</b> GMD019A00S	<b>Mod. 984e-ext</b>  <b>Rev. 07</b>
<b>Description</b> EPI-Baby IV Filter Neonatal 0.2 µm Luer Lock	

	<p><b>Maximum applicable pressure:</b> 8.0 bar  <b>Flow Rate:</b> ≥ 4 ml/min @ 80 cm (31.5 in) water head pressure  <b>Bubble point:</b> 3.7 ÷ 4.8 bar</p> <p><b>Chemical:</b>  <b>Compatibility to solvents:</b> Isopropyl Alcohol</p> <p><b>Biological:</b>  <b>Biocompatibility:</b> Compliant with ISO 10993-1</p> <p><b>Features:</b>  <b>Type of administration:</b> gravity / pressure  <b>Duration of the application:</b> up to 96 hours  <b>Filter:</b> EPI-Baby Neonatal 0.2 µm non-vented  <b>Filter pore size:</b> 0.2 µm  <b>Filter internal volume:</b> &lt; 0.35 ml  <b>Filter media:</b> hydrophilic PES membrane</p>
<b>PRODUCT SHELF LIFE</b>	5 years.
<b>STERILIZATION</b>	<p><b>Sterile:</b> Yes –Ethylene Oxide (EtO)  <b>Suitable for Resterilization:</b> No</p>
<b>APPLICABLE STANDARDS AND REGULATIONS</b>	<p><b>Product Certification:</b> CE mark</p> <p><b>Applicable Standards and Technical Regulations:</b></p> <p>EN 556-1 Sterilization of medical devices - Requirements for medical devices to be designated "STERILE" - Part 1: Requirements for terminally sterilized medical devices</p> <p>EN ISO 8536-4 Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed</p> <p>EN ISO 8536-9 Infusion equipment for medical use — Part 9: Fluid lines for single use with pressure infusion equipment</p> <p>EN ISO 8536-11 Infusion equipment for medical use — Part 11: Infusion filters for single use with pressure infusion equipment</p> <p>EN ISO 10993-1 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process</p> <p>EN ISO 10993-4 Biological evaluation of medical devices — Part 4: Selection of tests for interactions with blood</p> <p>EN ISO 10993-5 Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity</p> <p>EN ISO 10993-7 Biological evaluation of medical devices - Part 7: Ethylene oxide sterilization residuals</p> <p>EN ISO 10993-10 Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization</p> <p>EN ISO 10993-11 Biological evaluation of medical devices -- Part 11: Tests for systemic toxicity</p> <p>EN ISO 10993-18 Biological evaluation of medical devices - Part 18: Chemical characterization of materials</p> <p>EN ISO 10993-23 Biological evaluation of medical devices — Part 23: Tests for irritation</p> <p>EN ISO 11135 Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices</p> <p>EN ISO 11607-1 Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems</p> <p>EN ISO 11607-2 Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes</p> <p>EN ISO 13485 Medical devices - Quality management systems - Requirements for regulatory purposes</p> <p>EN ISO 14971 Medical devices - Application of risk management to medical devices</p> <p>EN ISO 15223-1 Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements</p> <p>EN ISO 20417 Medical devices - Information to be supplied by the manufacturer</p> <p>IEC 62366-1 Medical devices – Part 1: Application of usability engineering to medical devices</p>

<b>Product PN</b> GMD019A00S	<b>Mod. 984e-ext</b> <b>Rev. 07</b>
<b>Description</b> EPI-Baby IV Filter Neonatal 0.2 µm Luer Lock	

	EN ISO 80369-1 Small-bore connectors for liquids and gases in healthcare applications — Part 1: General requirements EN ISO 80369-7 Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications EN ISO 80369-20 Small-bore connectors for liquids and gases in healthcare applications — Part 20: Common test methods
--	--

<b>INSTRUCTIONS FOR USE</b>	Available in: English / Italian / German / French / Spanish
-----------------------------	--

<b>PACKAGING</b>	<b>Primary Packaging:</b> Devices are individually packed and label in medical paper pouches. Pouch Size: 100 X 145 mm  <b>Secondary Packaging:</b> Pouches are placed inside a microperforated bag.  <b>Tertiary Packaging:</b> Carton Box. Box Size: 60 x 40 x 20 cm Box Weight: 4 kg Devices per box: 400
------------------	---

<b>CERTIFICATIONS</b>	ISO 9001:2015 ISO 13485:2016 CE Certificate (2017/745/UE)
-----------------------	---

<b>DRAWING</b>		<table border="1"> <thead> <tr> <th>ID</th> <th>Description</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>Vented male rotating cap for female Luer Lock</td> </tr> <tr> <td>2</td> <td>Speedflow® filter Neonatal 0.2 µm non-vented</td> </tr> <tr> <td>3</td> <td>Vented protecting cap for rotating male Luer Lock</td> </tr> </tbody> </table>	ID	Description	1	Vented male rotating cap for female Luer Lock	2	Speedflow® filter Neonatal 0.2 µm non-vented	3	Vented protecting cap for rotating male Luer Lock
		ID	Description							
1	Vented male rotating cap for female Luer Lock									
2	Speedflow® filter Neonatal 0.2 µm non-vented									
3	Vented protecting cap for rotating male Luer Lock									

## REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED BY: (name/function/signature)	VERIFIED BY: (name/function/signature)	APPROVED BY: (name/function/signature)
26/02/2023	00	First issue	Barbara Palmieri RA Specialist 	Barbara Finessi QA Manage 	Luca Zanini VP Healthcare and Lifesience 