

Product PN GIS011D01S GIS011E01S

Mod. 984e-ext

Description HI-FLO IV Set Easydrop® double scale

Rev. 07

## HI-FLO IV Set with Regulator and y-injection

<b>PRODUCT DESCRIPTION</b>	Infusion set for infusion for intravascular applications. The device is sterile and is for gravity feed infusion only. The set has a vented IV drip chamber (20 drops/ml) and a male luer lock with ring and protective cap outlet, compliant with ISO 80369-7. The set is provided with a flow regulator Easydrop® double scale, to approximately regulate the flow rate. The set is provided with a y-injection site, for administration of additional drugs by means of a syringe using the same infusion line, and with a clamp. The approximate total length of the set is 2663 mm in GIS011D01S model and 1913 mm in the GIS011E01S model. The set includes the following components: <ul style="list-style-type: none"><li>▪ Drip Chamber (20 drops/ml);</li><li>▪ Clamp;</li><li>▪ Flow regulator Easydrop® double scale;</li><li>▪ Y-Injection point;</li><li>▪ Male luer lock with ring and protective cap outlet.</li></ul>
<b>CONFIGURATIONS</b>	GIS011D01S Total length 2663 mm GIS011E01S Total length 1913 mm
<b>MANUFACTURER NAME</b>	<b>GVS S.p.A.</b> Via Roma, 50 – 40069 Zola Predosa (BO) – Italy Phone : +39.051.6176311 – Fax: +39.051.6176200 email: <a href="mailto:gvs@gvs.it">gvs@gvs.it</a> – website: <a href="http://www.gvs.com">www.gvs.com</a>
<b>INTENDED USE / APPLICATION</b>	Infusion of Drugs (including chemotherapy), Anesthetics, Nutritional Solutions and Saline Solutions for intravascular applications. The set can be connected via Luer Locks to all commercially available infusion administration sets compliant with ISO 80369-7. The set is for gravity feed infusion. The set is a single-use device and has to be disposed after each therapy. The device should only be supervised and used by qualified healthcare personnel.
<b>CLASS OF THE PRODUCT</b>	Class Ia – sterile – single use Rule 2 Annex VIII 2017/745/UE
<b>REGISTRATION NUMBERS</b>	<b>Italian national database:</b> GIS011D01S 2439445 GIS011E01S 2439446
<b>EMDN</b>	A03010103 INFUSION CONTROLLERS WITH FLOW REGULATOR
<b>MATERIALS</b>	<b>Tubes:</b> PVC (DEHP Free) <b>Flow Regulator:</b> SEBS / ABS / Alkyl Polysiloxane <b>Drip Chamber:</b> ABS / PVC (DEHP Free) / LDPE / HDPE / PA6.6 / NY <b>Y-injection:</b> SEBS / MABS <b>Clamp:</b> PP / LDPE <b>Luer Lock:</b> MABS <b>Cap:</b> HDPE / PE  <b>Regulatory Compliance:</b> <ul style="list-style-type: none"><li>▪ Biocompatibility according to ISO 10993-1</li><li>▪ RoHS, directive 2011/65/EU</li></ul>



# PRODUCT SPECIFICATION

CE

Product PN GIS011D01S GIS011E01S

Mod. 984e-ext

Description HI-FLO IV Set Easydrop® double scale

Rev. 07

	<ul style="list-style-type: none"><li>DEHP plasticizer Free</li><li>Latex free</li><li>Reach 1907/2006/CE (hazardous substances regulation)</li><li>Reg. 1272/2008/CE (classification, labelling and packaging of substances and mixtures)</li></ul>
<b>PRODUCT CHARACTERISTICS</b>	<p><b>Physical/Mechanical:</b></p> <p><b>Approximate total length:</b> GIS011D01S: 2663 mm GIS011E01S: 1913 mm</p> <p><b>Total internal volume of the set:</b> GIS011D01S: 26.9 ml GIS011E01S: 29.7 ml</p> <p><b>Weight:</b> GIS011D01S: 41 g GIS011E01S: 36 g</p> <p><b>Input/output connectors:</b> Male luer lock outlet with ring and protective cap compliant with ISO 80369-7</p> <p><b>Operating temperature Range:</b> From 5 °C to 40 °C</p> <p><b>Storage temperature Range:</b> From 0 °C to 40 °C</p> <p><b>Biological:</b></p> <p><b>Biocompatibility:</b> Compliant with ISO 10993-1</p> <p><b>Features:</b></p> <p><b>Type of administration:</b> gravity</p> <p><b>Duration of the application:</b> N.A. The device must be disposed after each therapy.</p> <p><b>Filter:</b> No</p> <p><b>Flow Regulator:</b> Easydrop® double scale</p> <p><b>Drip Chamber:</b> Yes – 20 drops / ml</p> <p><b>Roller:</b> No</p> <p><b>Y-injection site:</b> Yes</p> <p><b>Clamp:</b> Yes</p> <p><b>Tubing:</b> PVC tube – inner Ø 3.0 mm / outer Ø 4.1 mm</p>
<b>PRODUCT SHELF LIFE</b>	5 years
<b>STERILIZATION</b>	<p><b>Sterile:</b> Yes – EtO</p> <p><b>Suitable for Sterilization/Re-sterilization:</b> No</p>
<b>APPLICABLE STANDARDS AND REGULATIONS</b>	<p><b>Product Certification:</b> CE mark</p> <p><b>Applicable Standards and Technical Regulations:</b></p> <p>EN 556-1 Sterilization of medical devices — Requirements for medical devices to be designated "STERILE" — Part 1: Requirements for terminally sterilized medical devices</p> <p>EN ISO 8536-4 Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed</p> <p>EN ISO 8536-13 Infusion equipment for medical use — Part 13: Graduated flow regulators for single use with fluid contact</p> <p>EN ISO 8536-14 Infusion equipment for medical use — Part 14: Clamps and flow regulators for transfusion and infusion equipment without fluid contact</p> <p>EN ISO 10993-1 Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process</p> <p>EN ISO 10993-4 Biological evaluation of medical devices — Part 4: Selection of tests for interactions with blood</p> <p>EN ISO 10993-5 Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity</p> <p>EN ISO 10993-7 Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals</p> <p>EN ISO 10993-10 Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization</p> <p>EN ISO 10993-11 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity</p> <p>EN ISO 10993-18 Biological evaluation of medical devices — Part 18: Chemical characterization of materials</p> <p>EN ISO 10993-23 Biological evaluation of medical devices — Part 23: Tests for irritation</p> <p>EN ISO 11135 Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices</p> <p>EN ISO 11607-1 Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems</p>

Product PN GIS011D01S GIS011E01S

Mod. 984e-ext

Description HI-FLO IV Set Easydrop® double scale

Rev. 07

	EN ISO 11607-2	Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes																		
	EN ISO 13485	Medical devices — Quality management systems — Requirements for regulatory purposes																		
	EN ISO 14971	Medical devices — Application of risk management to medical devices																		
	EN ISO 15223-1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements																		
	EN ISO 20417	Medical devices — Information to be supplied by the manufacturer																		
	IEC 62366-1	Medical devices — Part 1: Application of usability engineering to medical devices																		
	EN ISO 80369-1	Small-bore connectors for liquids and gases in healthcare applications — Part 1: General requirements																		
	EN ISO 80369-7	Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications																		
	EN ISO 80369-20	Small-bore connectors for liquids and gases in healthcare applications — Part 20: Common test methods																		
<b>INSTRUCTIONS FOR USE</b>	<b>Available languages:</b>	English / Italian / German / French / Spanish / Hungarian / Romanian																		
<b>PACKAGING</b>	<b>Primary Packaging:</b>	Devices are individually packed and label in medical paper bags. Bag Size: 150 X 250 mm																		
	<b>Secondary Packaging:</b>	Bags are placed inside a microperforated sack.																		
	<b>Tertiary Packaging:</b>	Carton Box. Box Size: 60 x 40 x 20 cm Box Weight: 4.3 kg Devices per box: 70																		
<b>CERTIFICATIONS</b>	ISO 9001:2015 ISO 13485:2016 CE Certificate (2017/745/UE)																			
<b>DRAWING</b>		<table border="1"> <thead> <tr> <th>ID</th> <th>Description</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>Drip Chamber 20 drops/ml</td> </tr> <tr> <td>2</td> <td>Tube 75 cm – inner Ø 3.0 mm / outer Ø 4.1 mm</td> </tr> <tr> <td>3</td> <td>Clamp</td> </tr> <tr> <td>4</td> <td>Flow Regulator Easydrop® double scale</td> </tr> <tr> <td>5</td> <td>Tube Ø 4.1 mm 'D' model: 150 cm 'E' model: 75 cm</td> </tr> <tr> <td>6</td> <td>Y-Injection site</td> </tr> <tr> <td>7</td> <td>Tube 15 cm – inner Ø 3.0 mm / outer Ø 4.1 mm</td> </tr> <tr> <td>8</td> <td>Male Luer Lock with ring and protective cap Ø 4.1 mm</td> </tr> </tbody> </table>	ID	Description	1	Drip Chamber 20 drops/ml	2	Tube 75 cm – inner Ø 3.0 mm / outer Ø 4.1 mm	3	Clamp	4	Flow Regulator Easydrop® double scale	5	Tube Ø 4.1 mm 'D' model: 150 cm 'E' model: 75 cm	6	Y-Injection site	7	Tube 15 cm – inner Ø 3.0 mm / outer Ø 4.1 mm	8	Male Luer Lock with ring and protective cap Ø 4.1 mm
ID	Description																			
1	Drip Chamber 20 drops/ml																			
2	Tube 75 cm – inner Ø 3.0 mm / outer Ø 4.1 mm																			
3	Clamp																			
4	Flow Regulator Easydrop® double scale																			
5	Tube Ø 4.1 mm 'D' model: 150 cm 'E' model: 75 cm																			
6	Y-Injection site																			
7	Tube 15 cm – inner Ø 3.0 mm / outer Ø 4.1 mm																			
8	Male Luer Lock with ring and protective cap Ø 4.1 mm																			



# PRODUCT SPECIFICATION

CE

Product PN GIS011D01S GIS011E01S

Mod. 984e-ext

Description HI-FLO IV Set Easydrop® double scale

Rev. 07

## REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED BY: (name/function/signature)	VERIFIED BY: (name/function/signature)	APPROVED BY: (name/function/signature)
14/02/2025	02	Added Hungarian and Romanian translations on labels and IFU	Barbara Palmieri RA Specialist 	Barbara Finessi QA Manage 	Luca Zanini VP Healthcare and Lifesience 