



Product PN	GIS013A01S	Mod. 984e-ext Rev. 07
Description	HI-FLO IV Set Hybrid Easydrop® double scale	

HI-FLO IV Set with Regulator and y-injection

PRODUCT DESCRIPTION	Infusion set for infusion therapies for intravascular applications. The device is sterile and is for gravity feed infusion only. The set has a vented IV drip chamber (20 drops/ml) and a male luer lock with rotating nut and female luer filter outlet, compliant with ISO 80369-7. The set is provided with a flow regulator Easydrop® double scale, to approximately regulate the flow rate. The set is provided with a y-injection site, for administration of additional drugs by means of a syringe using the same infusion line, and with a clamp. The approximate total length of the set is 2204 mm. The set includes the following components: ▪ Drip Chamber (20 drops/ml); ▪ Clamp; ▪ Flow regulator Easydrop® double scale; ▪ Y-Injection point; ▪ Male luer lock with rotating nut and female luer filter outlet.	
MANUFACTURER NAME	GVS S.p.A. Via Roma, 50 – 40069 Zola Predosa (BO) – Italy Phone : +39.051.6176311 – Fax: +39.051.6176200 email: gvs@gvs.it – website: www.gvs.com	
INTENDED USE / APPLICATION	Infusion of Drugs (including chemotherapy), Anesthetics, Nutritional Solutions and Saline Solutions for intravascular applications. The set can be connected via Luer Locks to all commercially available infusion administration sets compliant with ISO 80369-7. The set is for gravity feed infusion. The set is a single-use device and has to be disposed after each therapy. The device should only be supervised and used by qualified healthcare personnel.	
CLASS OF THE PRODUCT	Class Is – sterile – single use Rule 2 Annex VIII 2017/472/UE	
EMDN	A03010103	INFUSION CONTROLLERS WITH FLOW REGULATOR
MATERIALS	Tubes: PVC (DEHP Free) Flow Regulator: SEBS / ABS / Alkyl Polysiloxane Drip Chamber: ABS / PVC (DEHP Free) / LDPE / HDPE / PA6.6 / NY Y-injection: SEBS / MABS Clamp: PP / LDPE Luer Lock: MABS Caps: HDPE / PE Regulatory Compliance: ▪ Biocompatibility according to ISO 10993-1 ▪ Rohs directive 2011/65/EU ▪ DEHP plasticizer Free ▪ Latex free ▪ Reach 1907/2006/CE (hazardous substances regulation) ▪ Reg. 1272/2008/CE (classification, labelling and packaging of substances and mixtures)	




PRODUCT SPECIFICATION



Product PN	GIS013A01S	Mod. 984e-ext Rev. 07
Description	HI-FLO IV Set Hybrid Easydrop® double scale	

PRODUCT CHARACTERISTICS	Physical/Mechanical: Approximate total length: 2204 mm Total internal volume of the set: 29.5 ml Weight: 37.6 g Input/output connectors: Male luer lock outlet with rotating nut and female luer filter compliant with ISO 80369-7 Operating temperature Range: From 5 °C to 40 °C Storage temperature Range: From 0 °C to 40 °C	
	Biological: Biocompatibility: Compliant with ISO 10993-1	
PRODUCT SHELF LIFE	Features: Type of administration: gravity Duration of the application: N.A. The device must be disposed after each therapy. Filter: No Flow Regulator: Easydrop® double scale Drip Chamber: Yes – 20 drops / ml Roller: No Y-injection site: Yes Clamp: Yes Tubing: PVC tube – inner Ø 3.0 mm / outer Ø 4.1 mm	
	5 years	
STERILIZATION	Sterile:	Yes – EtO
	Suitable for Sterilization/Re-sterilization:	No
APPLICABLE STANDARDS AND REGULATIONS	Product Certification: CE mark	
	Applicable Standards and Technical Regulations: EN 556-1 Sterilization of medical devices — Requirements for medical devices to be designated "STERILE" — Part 1: Requirements for terminally sterilized medical devices EN ISO 8536-4 Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed EN ISO 8536-13 Infusion equipment for medical use — Part 13: Graduated flow regulators for single use with fluid contact EN ISO 8536-14 Infusion equipment for medical use — Part 14: Clamps and flow regulators for transfusion and infusion equipment without fluid contact EN ISO 10993-1 Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process EN ISO 10993-4 Biological evaluation of medical devices — Part 4: Selection of tests for interactions with blood EN ISO 10993-5 Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity EN ISO 10993-7 Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals EN ISO 10993-10 Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization EN ISO 10993-11 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity EN ISO 10993-18 Biological evaluation of medical devices — Part 18: Chemical characterization of materials EN ISO 10993-23 Biological evaluation of medical devices — Part 23: Tests for irritation EN ISO 11135 Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices EN ISO 11607-1 Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems EN ISO 11607-2 Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes EN ISO 13485 Medical devices — Quality management systems — Requirements for regulatory purposes EN ISO 14971 Medical devices — Application of risk management to medical devices	

Product PN GIS013A01S	Mod. 984e-ext Rev. 07
Description HI-FLO IV Set Hybrid Easydrop® double scale	

	EN ISO 15223-1 Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements EN ISO 20417 Medical devices — Information to be supplied by the manufacturer IEC 62366-1 Medical devices — Part 1: Application of usability engineering to medical devices EN ISO 80369-1 Small-bore connectors for liquids and gases in healthcare applications — Part 1: General requirements EN ISO 80369-7 Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications EN ISO 80369-20 Small-bore connectors for liquids and gases in healthcare applications — Part 20: Common test methods																				
INSTRUCTIONS FOR USE	Available languages: English / Italian / German / French / Spanish / Hungarian / Romanian																				
PACKAGING	Primary Packaging: Devices are individually packed and label in medical paper bags. Bag Size: 150 X 250 mm Secondary Packaging: Bags are placed inside a microperforated sack. Tertiary Packaging: Carton Box. Box Size: 60 x 40 x 20 cm Box Weight: 3.1 kg Devices per box: 50																				
CERTIFICATIONS	ISO 9001:2015 ISO 13485:2016 CE Certificate (2017/745/UE)																				
DRAWING	 <table border="1"> <thead> <tr> <th>ID</th><th>Description</th></tr> </thead> <tbody> <tr> <td>1</td><td>Drip Chamber 20 drops/ml</td></tr> <tr> <td>2</td><td>Tube 65 cm – inner Ø 3.0 mm / outer Ø 4.1 mm</td></tr> <tr> <td>3</td><td>Clamp</td></tr> <tr> <td>4</td><td>Flow Regulator Easydrop® double scale</td></tr> <tr> <td>5</td><td>Tube 112 cm – inner Ø 3.0 mm / outer Ø 4.1 mm</td></tr> <tr> <td>6</td><td>Y-Injection site</td></tr> <tr> <td>7</td><td>Tube 15 cm – inner Ø 3.0 mm / outer Ø 4.1 mm</td></tr> <tr> <td>8</td><td>Male Luer Lock with rotating nut Ø 4.1 mm</td></tr> <tr> <td>9</td><td>Female luer filter</td></tr> </tbody> </table>	ID	Description	1	Drip Chamber 20 drops/ml	2	Tube 65 cm – inner Ø 3.0 mm / outer Ø 4.1 mm	3	Clamp	4	Flow Regulator Easydrop® double scale	5	Tube 112 cm – inner Ø 3.0 mm / outer Ø 4.1 mm	6	Y-Injection site	7	Tube 15 cm – inner Ø 3.0 mm / outer Ø 4.1 mm	8	Male Luer Lock with rotating nut Ø 4.1 mm	9	Female luer filter
ID	Description																				
1	Drip Chamber 20 drops/ml																				
2	Tube 65 cm – inner Ø 3.0 mm / outer Ø 4.1 mm																				
3	Clamp																				
4	Flow Regulator Easydrop® double scale																				
5	Tube 112 cm – inner Ø 3.0 mm / outer Ø 4.1 mm																				
6	Y-Injection site																				
7	Tube 15 cm – inner Ø 3.0 mm / outer Ø 4.1 mm																				
8	Male Luer Lock with rotating nut Ø 4.1 mm																				
9	Female luer filter																				



PRODUCT SPECIFICATION



Product PN	GIS013A01S	Mod. 984e-ext Rev. 07
Description	HI-FLO IV Set Hybrid Easydrop® double scale	

REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED BY: (name/function/signature)	VERIFIED BY: (name/function/signature)	APPROVED BY: (name/function/signature)
14/02/2025	01	Added Hungarian and Romanian translations on labels and IFU	Barbara Palmieri RA Specialist 	Barbara Finessi QA Manage 	Luca Zanini VP Healthcare and Lifesience 