

## UK DECLARATION OF CONFORMITY



## The manufacturer

GVS Filter Technology UK Limited NFC House, Vickers Industrial Estate Mellishaw Lane, Morecambe, LA3 3EN UK

Here declares under sole responsibility that the personal protective equipment (PPE) in category III:

Part Number: Description:

F31000 NR Segre Folded P3 Disposable Medical Facemask

**Protection:** 

FFP3 NR / Type IIR

It is produced in compliance with the essential health and safety requirements according to Regulation (EU) 2016/425 of the European Parliament and of Council of 09 March 2016 on personal protective equipment regulation as adopted into UK law and amended.

EN 149:2001+A1:2009 Respiratory protective devices. Filtering half masks to protect against particles. Requirements, testing, marking

Designated Standard EN 14683 Medical Facemasks - Requirements and methods

Is identical to the PPE which is the subject of UKCA Type certificate no.

**UKCA 746495** 

Issued by Approved Body no. 0086: BSI Assurance UK Ltd., Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 8PP, United Kingdom

The PPE listed above is manufactured under a system which satisfies the requirements of Module D (Annex VIII Cert UKCA 746494) & Module B of the same regulation as adopted into UK law and amended, which were Issued by Approved Body no. 0086: BSI Assurance UK Ltd., Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 8PP, United Kingdom

GVS Filter Technology UK Limited

08 July 2022

Andrew Pearce, Quality & Regulatory Manager



GVS Filter Technology UK Ltd.

NFC House - Mellishaw Lane - Morecambe, Lancashire United Kingdom LA3 3EN

Tel: +44 (0)1524 847600 - www.gvs.com

VAT no. 582426236 - Registered in England No. 2647901



ISO 9001
Outly
Management
Systems
CERTH ED

FM 713569

MOD 3015 Rev 0 Pg. 1